

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100223\
 Data File : PR063807.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2023 15:58
 Operator : YP\AJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12423391

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 22:32:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.666	2.880	131.8E6	71956009	20.314	21.048
2)	SA Decachlor...	9.665	8.172	134.3E6	103.4E6	36.268	34.582
Target Compounds							
16)	L4 AR-1242-1	4.906	4.004	66738303	38013956	400.317	398.193
17)	L4 AR-1242-2	4.928	4.021	96856905	58089117	405.915	406.798
18)	L4 AR-1242-3	4.992	4.201	59219824	29585208	417.081	403.878
19)	L4 AR-1242-4	5.096	4.294	46928055	27346484	400.398	399.820
20)	L4 AR-1242-5	5.892	4.844	45434317	35775604	391.807	408.978

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100223\
Data File : PR063807.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2023 15:58
Operator : YP\AJ
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12423391

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 22:32:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

