

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100418\
 Data File : PR032659.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2018 05:19
 Operator : SM\SJ
 Sample : J5183-21
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 JLC38

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 09:36:53 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 28 17:37:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.636	485.7E6	904.7E6	24.425	26.046
2) SA Decachlor...	10.225	8.481	777.3E6	530.6E6	23.390	20.177
Target Compounds						
26) L6 AR-1254-1	6.510	5.457	986.5E6	2417.8E6	810.177	547.284 #
27) L6 AR-1254-2	6.728	5.600	1460.4E6	1862.3E6	772.796	470.642 #
28) L6 AR-1254-3	7.093	5.995	1532.5E6	3579.8E6	744.707	576.029
29) L6 AR-1254-4	7.377	6.218	1050.5E6	1380.6E6	649.965	290.206 #
30) L6 AR-1254-5	7.793	6.624	1115.4E6	2054.9E6	644.202	476.511 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100418\
Data File : PR032659.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2018 05:19
Operator : SM\SJ
Sample : J5183-21
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
JLC38

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 09:36:53 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 28 17:37:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

