

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100623\
 Data File : PR063861.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2023 16:42
 Operator : AJ\MA
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR16603396

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/09/2023
 Supervised By :Ankita Jodhani 10/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 16:44:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.666	2.881	131.5E6	70938575	20.256	20.750
2)	SA Decachlor...	9.662	8.171	141.4E6	112.7E6	38.192	37.695
Target Compounds							
3)	L1 AR-1016-1	4.905	4.005	78080416	44418426	399.668	394.586
4)	L1 AR-1016-2	4.927	4.021	115.8E6	68349462	410.882	402.311
5)	L1 AR-1016-3	4.992	4.201	71339197	34957454	424.609	402.166
6)	L1 AR-1016-4	5.095	4.254	57494372	28228120	418.900	417.391
7)	L1 AR-1016-5	5.415	4.471	57726602	35046288	432.573	404.687
31)	L7 AR-1260-1	6.638	5.570	101.5E6	74008075	422.389	417.159
32)	L7 AR-1260-2	6.924	5.778	114.8E6	89008476	418.434	418.139
33)	L7 AR-1260-3	7.318	5.935	80733840	81725580	422.202	410.309m
34)	L7 AR-1260-4	7.561	6.445	87208393	67843174	398.833	414.315
35)	L7 AR-1260-5	7.903	6.715	165.1E6	159.4E6	381.530	402.609

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100623\
Data File : PR063861.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2023 16:42
Operator : AJ\MA
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR16603396

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2023
Supervised By :Ankita Jodhani 10/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 16:44:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

