

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100623\
 Data File : PR063858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2023 15:44
 Operator : AJ\MA
 Sample : 04744-08MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 BH580MS

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 15:57:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 10/09/2023
 Supervised By :Ankita Jodhani 10/09/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
 R.T. Response Conc
 6.64 107643375 448.05
 6.92 119949576 437.29
 7.32 70054143 366.35
 7.56 83465083 381.71
 7.90 149639115 345.77

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.57 78691799 443.56
 5.78 93849517 440.88
 5.94 86039034 431.96
 6.45 61620329 376.31
 6.71 147537250 372.71

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100623\
Data File : PR063858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2023 15:44
Operator : AJ\MA
Sample : 04744-08MS
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
BH580MS

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 15:57:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Reviewed By :Yogesh Patel 10/09/2023
Supervised By :Ankita Jodhani 10/09/2023

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
6.64 107643375 448.05
6.92 119949576 437.29
7.32 70054143 366.35
7.56 83465083 381.71
7.90 149639115 345.77

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.57 78691799 443.56
5.78 93849517 440.88
5.94 88067717 442.15
6.45 61620329 376.31
6.71 147537250 372.71