

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101023\
 Data File : PR063919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 23:23
 Operator : AJ\MA
 Sample : 04699-07
 Misc : AR1221 LOD 40 PPB
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 06:50:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 10 14:16:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.737	3.026	135.1E6	86451616	21.409	22.117
2) SA Decachlor...	9.757	8.428	108.4E6	67217767	22.961	22.049
Target Compounds						
8) L2 AR-1221-1	3.964	3.259	3154168	2078480	42.598	43.066
9) L2 AR-1221-2	4.053	3.343	3628306	2493842	68.923	67.341
10) L2 AR-1221-3	4.132	3.426	7688338	5198704	45.706	46.866

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101023\
Data File : PR063919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 23:23
Operator : AJ\MA
Sample : 04699-07
Misc : AR1221 LOD 40 PPB
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 06:50:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101023.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 10 14:16:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

