

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101218\
 Data File : PR032812.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2018 01:50
 Operator : SM\SJ
 Sample : MDL-AR1232-W-6
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1232-W-6

Manual Integrations
 APPROVED

Sohil
 10/15/2018 11:39:46 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 03:52:01 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101018CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 12 03:06:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.478f	3.753	309.5E6	855.1E6	20.246m	19.345
2) SA Decachlor...	10.105	8.747	507.2E6	1888.5E6	46.960m	54.002
Target Compounds						
11) L3 AR-1232-1	4.841f	4.129	10309399	22805887	33.769	26.609
12) L3 AR-1232-2	5.362f	4.853	5507249	32947372	33.007	31.210m
13) L3 AR-1232-3	5.707	5.029	8642326	16732353	23.911	30.733 #
14) L3 AR-1232-4	5.894	5.109	4238154	17248122	23.140	37.011 #
15) L3 AR-1232-5	5.976	5.280	2371624	19461239	17.831	35.615m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101218\
Data File : PR032812.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2018 01:50
Operator : SM\SJ
Sample : MDL-AR1232-W-6
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1232-W-6

Manual Integrations
APPROVED

Sohil
10/15/2018 11:39:46 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 03:52:01 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101018CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 12 03:06:23 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

