

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101222\
 Data File : PR057267.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2022 02:46
 Operator : AJ\MA
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1262ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/13/2022
 Supervised By :Ankita Jodhani 10/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 13 04:31:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 13 04:31:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.612	2.882	161.4E6	70274488	50.000	50.000
2) SA Decachlor...	9.489	8.077	122.0E6	80430785	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	7.218	6.109	91180501	48345442	500.000	500.000
37) L8 AR-1262-2	7.800	6.380	158.7E6	40976332	500.000	500.000
38) L8 AR-1262-3	8.105	6.947	106.9E6	40116979	500.000	500.000
39) L8 AR-1262-4	8.190	7.015	56471519	73244896	501.762m	500.000
40) L8 AR-1262-5	8.806	7.552	55965549	36342075	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101222\
Data File : PR057267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2022 02:46
Operator : AJ\MA
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2022
Supervised By :Ankita Jodhani 10/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 13 04:31:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 13 04:31:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.612 min
Delta R.T.: 0.000 min
Response: 161426991
Conc: 50.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2022
Supervised By :Ankita Jodhani 10/13/2022




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: 0.000 min
Response: 70274488
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.489 min
Delta R.T.: 0.000 min
Response: 121996789
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.077 min
Delta R.T.: 0.000 min
Response: 80430785
Conc: 50.00 ng/ml




#36 AR-1262-1

R.T.: 7.218 min
Delta R.T.: 0.000 min
Response: 91180501
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2022
Supervised By :Ankita Jodhani 10/13/2022




#36 AR-1262-1

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 48345442
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 7.800 min
Delta R.T.: 0.000 min
Response: 158722643
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 6.380 min
Delta R.T.: 0.000 min
Response: 40976332
Conc: 500.00 ng/ml




#38 AR-1262-3

R.T.: 8.105 min
Delta R.T.: 0.000 min
Response: 106901119
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2022
Supervised By :Ankita Jodhani 10/13/2022




#38 AR-1262-3

R.T.: 6.947 min
Delta R.T.: 0.000 min
Response: 40116979
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 8.190 min
Delta R.T.: 0.000 min
Response: 56471519
Conc: 501.76 ng/ml m



#39 AR-1262-4

R.T.: 7.015 min
Delta R.T.: 0.000 min
Response: 73244896
Conc: 500.00 ng/ml




#40 AR-1262-5

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 55965549
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1262ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2022
Supervised By :Ankita Jodhani 10/13/2022



#40 AR-1262-5

R.T.: 7.552 min
Delta R.T.: 0.000 min
Response: 36342075
Conc: 500.00 ng/ml