

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101223\
 Data File : PR064006.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2023 20:22
 Operator : AJ\MA
 Sample : AR1254ICC100
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 13 03:14:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 13 03:13:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.736	3.025	37815896	23305031	5.582	5.703
2) SA Decachlor...	9.758	8.425	53997932	35106422	11.609	11.803
Target Compounds						
26) L6 AR-1254-1	5.897	5.034	29579755	27394979	120.837	122.446
27) L6 AR-1254-2	6.136	5.195	44380994	23494947	119.403	121.889
28) L6 AR-1254-3	6.532	5.628	44464176	36044567	116.743	117.282
29) L6 AR-1254-4	6.846	5.880	30135496	21857407	114.233	119.585
30) L6 AR-1254-5	7.304	6.335	32978144	30980577	113.729	115.425



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101223\
Data File : PR064006.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2023 20:22
Operator : AJ\MA
Sample : AR1254ICC100
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 13 03:14:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 13 03:13:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 5.897 min
Delta R.T.: 0.000 min
Response: 29579755
Conc: 120.84 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 5.034 min
Delta R.T.: 0.001 min
Response: 27394979
Conc: 122.45 ng/ml




#27 AR-1254-2

R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 44380994
Conc: 119.40 ng/ml



#27 AR-1254-2


R.T.: 5.195 min
Delta R.T.: 0.000 min
Response: 23494947
Conc: 121.89 ng/ml



#28 AR-1254-3


R.T.: 6.532 min
Delta R.T.: 0.000 min
Response: 44464176
Conc: 116.74 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 36044567
Conc: 117.28 ng/ml




#29 AR-1254-4

R.T.: 6.846 min
Delta R.T.: 0.000 min
Response: 30135496
Conc: 114.23 ng/ml



#29 AR-1254-4


R.T.: 5.880 min
Delta R.T.: 0.000 min
Response: 21857407
Conc: 119.58 ng/ml



#30 AR-1254-5

R.T.: 7.304 min
Delta R.T.: 0.000 min
Response: 32978144
Conc: 113.73 ng/ml

Instrument :
ECD_R
ClientSampleId :



#30 AR-1254-5

R.T.: 6.335 min
Delta R.T.: 0.000 min
Response: 30980577
Conc: 115.42 ng/ml