

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101320\
Data File : PR047621.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2020 09:55
Operator : AJ\MA
Sample : L4267-09
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampled :
BEQC0

Manual Integrations
APPROVED

Ankita
10/16/2020 9:54:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 01:13:38 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR100920CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 10 06:29:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)
R.T. Response Conc
6.78 314251119 4780.80
6.99 588453013 5718.52
7.37 730116847 6758.94
7.65 637260170 7262.08
8.07 938298685 9738.39
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.79 797916565 3363.15
5.93 794859408 3486.86
6.34 2326610709 5193.69
6.57 2250874880 5088.79
6.99 4418543218 7430.43

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101320\
 Data File : PR047621.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2020 09:55
 Operator : AJ\MA
 Sample : L4267-09
 Misc :
 ALS Vial : 10 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampled :
 BEQC0

Manual Integrations
 APPROVED

Ankita
 10/16/2020 9:54:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 01:13:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR100920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 10 06:29:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
 R.T. Response Conc
 6.78 314251119 4780.80
 6.99 588453013 5718.52
 7.37 730116847 6758.94
 7.65 637260170 7262.08
 8.07 938298685 9738.39
 (26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 5.79 797916565 3363.15
 5.93 865386756 3796.25
 6.34 2326610709 5193.69
 6.57 2250874880 5088.79
 6.99 4418543218 7430.43

QEdit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101320\
 Data File : PR047621.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2020 09:55
 Operator : AJ\MA
 Sample : L4267-09
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BEQC0

Manual Integrations
 APPROVED

Ankita
 10/16/2020 9:54:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 01:13:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR100920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 10 06:29:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.736	3.886	17477285	42152541	14.553	12.020
2)	SA Decachlor...	10.682	8.973	52459533	304.2E6	29.440	25.218
Target Compounds							
21)	L5 AR-1248-1	5.917	4.977	47081435	93436991	1389.717	1189.030
22)	L5 AR-1248-2	6.191	5.215	201.3E6	331.1E6	4062.027	3498.553
23)	L5 AR-1248-3	6.398	5.257	92724930	285.9E6	1644.820	2459.901 #
24)	L5 AR-1248-4	6.802	5.431	78801339	193.3E6	1183.927	1258.913
25)	L5 AR-1248-5	6.845	5.826	113.8E6	126.3E6	1769.062	586.737 #
26)	L6 AR-1254-1	6.777	5.785	314.3E6	797.9E6	4780.798	3363.150 #
27)	L6 AR-1254-2	6.994	5.932	588.5E6	865.4E6	5718.522	3796.250m#
28)	L6 AR-1254-3	7.365	6.340	730.1E6	2326.6E6	6758.944	5193.689
29)	L6 AR-1254-4	7.651	6.568	637.3E6	2250.9E6	7262.083	5088.793 #
30)	L6 AR-1254-5	8.071	6.987	938.3E6	4418.5E6	9738.390	7430.425

AJ
 10/20/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.