

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101322\
 Data File : PR057319.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2022 21:20
 Operator : AJ\MA
 Sample : N5075-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 CLI-0637

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/14/2022
 Supervised By :Ankita Jodhani 10/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 08:46:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 13 04:58:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.612	2.883	20614891	8418269	5.953m	5.975m
2) SA Decachlor...	9.489	8.078	15708565	9878091	5.968	5.357

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101322\
Data File : PR057319.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2022 21:20
Operator : AJ\MA
Sample : N5075-01
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
CLI-0637


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/14/2022
Supervised By :Ankita Jodhani 10/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 08:46:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 13 04:58:31 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.612 min
Delta R.T.: 0.000 min
Response: 20614891
Conc: 5.95 ng/ml

Instrument :
ECD_R
ClientSampleId :
CLI-0637


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/14/2022
Supervised By :Ankita Jodhani 10/14/2022




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 8418269
Conc: 5.98 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.489 min
Delta R.T.: -0.003 min
Response: 15708565
Conc: 5.97 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.078 min
Delta R.T.: -0.003 min
Response: 9878091
Conc: 5.36 ng/ml