

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101421\
 Data File : PR052280.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2021 02:45
 Operator : AJ\MA
 Sample : AR1268ICC200
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 09:29:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101421CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 15 09:28:44 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.689	3.855	39331903	28348980	10.826	10.767
2) SA Decachlor...	10.632	8.875	115.6E6	113.4E6	21.286	21.258
Target Compounds						
41) L9 AR-1268-1	9.013	7.746	94503852	82033035	220.501	210.350
42) L9 AR-1268-2	9.115	7.810	85059681	79273998	216.698	209.220
43) L9 AR-1268-3	9.358	8.017	73597323	69992645	211.725	210.718
44) L9 AR-1268-4	9.825	8.308	35590647	31483442	220.836	217.032
45) L9 AR-1268-5	10.269	8.610	230.2E6	238.7E6	205.669	205.053



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101421\
Data File : PR052280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2021 02:45
Operator : AJ\MA
Sample : AR1268ICC200
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 09:29:25 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101421CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 15 09:28:44 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#41 AR-1268-1


R.T.: 9.013 min
Delta R.T.: 0.000 min
Response: 94503852
Conc: 220.50 ng/ml

Instrument :
ECD_R
ClientSampleId :




#41 AR-1268-1

R.T.: 7.746 min
Delta R.T.: 0.000 min
Response: 82033035
Conc: 210.35 ng/ml




#42 AR-1268-2

R.T.: 9.115 min
Delta R.T.: 0.000 min
Response: 85059681
Conc: 216.70 ng/ml



#42 AR-1268-2


R.T.: 7.810 min
Delta R.T.: 0.000 min
Response: 79273998
Conc: 209.22 ng/ml



#43 AR-1268-3


R.T.: 9.358 min
Delta R.T.: 0.000 min
Response: 73597323
Conc: 211.72 ng/ml

Instrument :
ECD_R
ClientSampleId :




#43 AR-1268-3

R.T.: 8.017 min
Delta R.T.: 0.000 min
Response: 69992645
Conc: 210.72 ng/ml




#44 AR-1268-4

R.T.: 9.825 min
Delta R.T.: -0.001 min
Response: 35590647
Conc: 220.84 ng/ml



#44 AR-1268-4


R.T.: 8.308 min
Delta R.T.: 0.000 min
Response: 31483442
Conc: 217.03 ng/ml



#45 AR-1268-5

R.T.: 10.269 min
Delta R.T.: 0.000 min
Response: 230183622
Conc: 205.67 ng/ml

Instrument :
ECD_R
ClientSampleId :



#45 AR-1268-5

R.T.: 8.610 min
Delta R.T.: 0.000 min
Response: 238687704
Conc: 205.05 ng/ml