

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101421\
Data File : PR052276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2021 01:43
Operator : AJ\MA
Sample : AR1262ICC400
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_R
LabSampleId :
AR1262ICC400

Manual Integrations
APPROVED

mohammad
10/18/2021 2:43:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 08:37:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101421CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 15 08:36:21 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(36) AR-1262-1 (L8)
R.T. Response Conc
8.10 81636897 400.00
8.68 149234756 400.00
9.01 67812404 400.00
9.11 80068689 664.77
9.83 58592376 400.00

(36) AR-1262-1 #2 (L8)
R.T. Response Conc
6.76 50290773 400.00
7.46 136195254 400.00
7.75 55706841 400.00
7.81 107373810 400.00
8.31 55192404 400.00

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101421\
Data File : PR052276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2021 01:43
Operator : AJ\MA
Sample : AR1262ICC400
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_R
LabSampleId :
AR1262ICC400

Manual Integrations
APPROVED

mohammad
10/18/2021 2:43:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 08:37:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101421CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 15 08:36:21 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(36) AR-1262-1 (L8)
R.T. Response Conc
8.10 81636897 400.00
8.68 149234756 400.00
9.01 67812404 400.00
9.11 46622108 387.08
9.83 58592376 400.00

(36) AR-1262-1 #2 (L8)
R.T. Response Conc
6.76 50290773 400.00
7.46 136195254 400.00
7.75 55706841 400.00
7.81 107373810 400.00
8.31 55192404 400.00

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101421\
 Data File : PR052276.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2021 01:43
 Operator : AJ\MA
 Sample : AR1262ICC400
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1262ICC400

Manual Integrations
 APPROVED

mohammad
 10/18/2021 2:43:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 08:37:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101421CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 15 08:36:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.689	3.856	75939305	55015042	20.000	20.000
2) SA Decachlor...	10.632	8.875	136.6E6	131.3E6	40.000	40.000
Target Compounds						
36) L8 AR-1262-1	8.103	6.757	81636897	50290773	400.000	400.000
37) L8 AR-1262-2	8.683	7.462	149.2E6	136.2E6	400.000	400.000
38) L8 AR-1262-3	9.013	7.746	67812404	55706841	400.000	400.000
39) L8 AR-1262-4	9.111	7.809	46622108	107.4E6	387.082m	400.000
40) L8 AR-1262-5	9.825	8.307	58592376	55192404	400.000	400.000

10/18/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.