

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101620\
 Data File : PR047760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2020 13:35
 Operator : AJ\MA
 Sample : L4402-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AB8

Manual Integrations
 APPROVED

Ankita
 10/19/2020 2:06:42 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 08:49:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR100920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 10 06:29:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.738	3.888	22517860	45824653	18.750	13.067 #
2) SA Decachlor...	10.677	8.972	81233073	548.4E6	45.588	45.464
Target Compounds						
21) L5 AR-1248-1	5.919	4.980	833.1E6	1295.8E6	24589.541m	16489.298m#
22) L5 AR-1248-2	6.190	5.215	5963.9E6	7515.8E6	120318.815	79425.549 #
23) L5 AR-1248-3	6.398	5.258	6834.2E6	9482.0E6	121229.946	81592.591 #
24) L5 AR-1248-4	6.802	5.432	4595.5E6	13364.6E6	69043.947	87034.925 #
25) L5 AR-1248-5	6.842	5.827	10958.1E6	25416.2E6	170417.929	118105.404 #
31) L7 AR-1260-1	7.527	6.470	9113.3E6	27486.2E6	107189.710	110412.382
32) L7 AR-1260-2	7.782	6.668	11404.7E6	44632.3E6	109504.741	75902.862 #
33) L7 AR-1260-3	8.143	6.818	6084.8E6	36766.9E6	75152.657	91108.071
34) L7 AR-1260-4	8.381	7.293	9036.3E6	33589.8E6	92168.760	75048.458
35) L7 AR-1260-5	8.724	7.546	15101.7E6	75223.8E6	77025.870	38538.614 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101620\
Data File : PR047760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2020 13:35
Operator : AJ\MA
Sample : L4402-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
C0AB8

Manual Integrations
APPROVED

Ankita
10/19/2020 2:06:42 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 08:49:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR100920CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 10 06:29:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

