

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101719\
 Data File : PR042296.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2019 14:00
 Operator : SM\AJ
 Sample : K5111-07
 Misc : AR1660 LOD 25 PPB
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2019

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:31:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 17 06:40:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.724	3.984	90852238	347.7E6	19.077	20.420
2)	SA Decachlor...	10.650	9.096	71080069	212.2E6	20.330	20.903
Target Compounds							
3)	L1 AR-1016-1	5.903	5.082	4637934	17016240	29.672	31.907
4)	L1 AR-1016-2	5.926	5.102	6039127	22297802	27.132	30.338
5)	L1 AR-1016-3	5.989	5.281	3832432	9222301	29.011	28.411
6)	L1 AR-1016-4	6.090	5.320	3212844	7556413	29.686	28.933
7)	L1 AR-1016-5	6.383	5.537	2663534	9780744	24.623	33.000 #
9)	L2 AR-1221-2	0.000	4.289	0	6940366	N.D.	54.899 #
10)	L2 AR-1221-3	0.000	4.362	0	8199153	N.D.	20.370 #
11)	L3 AR-1232-1	0.000	4.362	0	8199153	N.D.	14.119 #
12)	L3 AR-1232-2	0.000	5.102	0	22297802	N.D.	39.613 #
13)	L3 AR-1232-3	5.926	5.281	6039127	9222301	36.043	36.462
14)	L3 AR-1232-4	6.090	5.364	3212844	7566718	39.201	35.909
15)	L3 AR-1232-5	0.000	5.537	0	9780744	N.D.	46.629 #
16)	L4 AR-1242-1	5.903	5.082	4637934	17016240	39.552	41.318
17)	L4 AR-1242-2	5.926	5.102	6039127	22297802	35.793	40.046
18)	L4 AR-1242-3	5.989	5.281	3832432	9222301	38.128	34.903
19)	L4 AR-1242-4	6.090	5.364	3212844	7566718	38.499	31.370
20)	L4 AR-1242-5	6.761	5.892	2570146	5892033	26.822	22.151
21)	L5 AR-1248-1	5.903	5.082	4637934	17016240	50.110	51.674
22)	L5 AR-1248-2	0.000	5.320	0	7556413	N.D.	21.553 #
23)	L5 AR-1248-3	6.383	5.364	2663534	7566718	18.528	21.554
24)	L5 AR-1248-4	6.761	5.537	2570146	9780744	15.107	25.737 #
25)	L5 AR-1248-5	6.761	5.892	2570146	5892033	16.006	15.828
26)	L6 AR-1254-1	6.761	5.892	2570146	5892033	14.308	10.250 #
27)	L6 AR-1254-2	0.000	6.038	0	7227196	N.D.	13.212 #
28)	L6 AR-1254-3	0.000	6.459	0	24680092	N.D.	25.668 #
30)	L6 AR-1254-5	8.053	7.092	8003334	24752861	34.511	30.141
31)	L7 AR-1260-1	7.511	6.576	6689797	19615371	32.161	33.913
32)	L7 AR-1260-2	7.766	6.761	7274801	27243370	29.627	34.512
33)	L7 AR-1260-3	8.127	6.918	5122861	20245209	27.456	29.769
34)	L7 AR-1260-4	8.367	7.391	7159283	15639764	33.986	28.544
35)	L7 AR-1260-5	8.705	7.631	11673344	43564173	27.784	29.565
36)	L8 AR-1262-1	8.127	7.184	5122861	10498421	17.483	25.745 #
37)	L8 AR-1262-2	8.705	7.631	11673344	43564173	21.464	23.530
38)	L8 AR-1262-3	9.058	7.917	7658686	8474136	22.360	12.202 #
39)	L8 AR-1262-4	0.000	7.980	0	30111292	N.D.	24.673 #
40)	L8 AR-1262-5	0.000	8.498	0	9779889	N.D.	18.534 #
41)	L9 AR-1268-1	9.058	7.917	7658686	8474136	13.974	4.513 #
42)	L9 AR-1268-2	0.000	7.980	0	30111292	N.D.	18.042 #
43)	L9 AR-1268-3	0.000	8.196	0	7094766	N.D.	5.046 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101719\
 Data File : PR042296.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2019 14:00
 Operator : SM\AJ
 Sample : K5111-07
 Misc : AR1660 LOD 25 PPB
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2019

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:31:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 17 06:40:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	0.000	8.498	0	9779889	N.D.	18.069 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101719\
Data File : PR042296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2019 14:00
Operator : SM\AJ
Sample : K5111-07
Misc : AR1660 LOD 25 PPB
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:31:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 17 06:40:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 90852238
Conc: 19.08 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#1 Tetrachloro-m-xylene

R.T.: 3.984 min
Delta R.T.: 0.000 min
Response: 347671939
Conc: 20.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.650 min
Delta R.T.: 0.002 min
Response: 71080069
Conc: 20.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.096 min
Delta R.T.: -0.002 min
Response: 212181149
Conc: 20.90 ng/ml



#3 AR-1016-1


R.T.: 5.903 min
Delta R.T.: 0.001 min
Response: 4637934
Conc: 29.67 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#3 AR-1016-1

R.T.: 5.082 min
Delta R.T.: 0.000 min
Response: 17016240
Conc: 31.91 ng/ml




#4 AR-1016-2

R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 6039127
Conc: 27.13 ng/ml



#4 AR-1016-2


R.T.: 5.102 min
Delta R.T.: 0.000 min
Response: 22297802
Conc: 30.34 ng/ml



#5 AR-1016-3


R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 3832432
Conc: 29.01 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#5 AR-1016-3

R.T.: 5.281 min
Delta R.T.: 0.000 min
Response: 9222301
Conc: 28.41 ng/ml




#6 AR-1016-4

R.T.: 6.090 min
Delta R.T.: 0.000 min
Response: 3212844
Conc: 29.69 ng/ml



#6 AR-1016-4


R.T.: 5.320 min
Delta R.T.: -0.001 min
Response: 7556413
Conc: 28.93 ng/ml



#7 AR-1016-5


R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 2663534
Conc: 24.62 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#7 AR-1016-5

R.T.: 5.537 min
Delta R.T.: -0.001 min
Response: 9780744
Conc: 33.00 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 5.020 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.289 min
Delta R.T.: 0.004 min
Response: 6940366
Conc: 54.90 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 5.099 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#10 AR-1221-3

R.T.: 4.362 min
Delta R.T.: 0.000 min
Response: 8199153
Conc: 20.37 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 5.099 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.362 min
Delta R.T.: 0.000 min
Response: 8199153
Conc: 14.12 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T.: 5.634 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#12 AR-1232-2

R.T.: 5.102 min
Delta R.T.: 0.000 min
Response: 22297802
Conc: 39.61 ng/ml




#13 AR-1232-3

R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 6039127
Conc: 36.04 ng/ml



#13 AR-1232-3


R.T.: 5.281 min
Delta R.T.: 0.000 min
Response: 9222301
Conc: 36.46 ng/ml



#14 AR-1232-4


R.T.: 6.090 min
Delta R.T.: 0.000 min
Response: 3212844
Conc: 39.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#14 AR-1232-4

R.T.: 5.364 min
Delta R.T.: 0.000 min
Response: 7566718
Conc: 35.91 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 6.176 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.537 min
Delta R.T.: 0.000 min
Response: 9780744
Conc: 46.63 ng/ml



#16 AR-1242-1


R.T.: 5.903 min
Delta R.T.: 0.002 min
Response: 4637934
Conc: 39.55 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#16 AR-1242-1

R.T.: 5.082 min
Delta R.T.: 0.000 min
Response: 17016240
Conc: 41.32 ng/ml




#17 AR-1242-2

R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 6039127
Conc: 35.79 ng/ml



#17 AR-1242-2


R.T.: 5.102 min
Delta R.T.: 0.000 min
Response: 22297802
Conc: 40.05 ng/ml



#18 AR-1242-3


R.T.: 5.989 min
Delta R.T.: 0.001 min
Response: 3832432
Conc: 38.13 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#18 AR-1242-3

R.T.: 5.281 min
Delta R.T.: 0.000 min
Response: 9222301
Conc: 34.90 ng/ml




#19 AR-1242-4

R.T.: 6.090 min
Delta R.T.: 0.001 min
Response: 3212844
Conc: 38.50 ng/ml



#19 AR-1242-4


R.T.: 5.364 min
Delta R.T.: 0.000 min
Response: 7566718
Conc: 31.37 ng/ml



#20 AR-1242-5


R.T.: 6.761 min
Delta R.T.: -0.065 min
Response: 2570146
Conc: 26.82 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#20 AR-1242-5

R.T.: 5.892 min
Delta R.T.: 0.002 min
Response: 5892033
Conc: 22.15 ng/ml




#21 AR-1248-1

R.T.: 5.903 min
Delta R.T.: 0.001 min
Response: 4637934
Conc: 50.11 ng/ml



#21 AR-1248-1


R.T.: 5.082 min
Delta R.T.: 0.000 min
Response: 17016240
Conc: 51.67 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 6.175 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#22 AR-1248-2

R.T.: 5.320 min
Delta R.T.: -0.001 min
Response: 7556413
Conc: 21.55 ng/ml




#23 AR-1248-3

R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 2663534
Conc: 18.53 ng/ml



#23 AR-1248-3


R.T.: 5.364 min
Delta R.T.: 0.000 min
Response: 7566718
Conc: 21.55 ng/ml



#24 AR-1248-4


R.T.: 6.761 min
Delta R.T.: -0.026 min
Response: 2570146
Conc: 15.11 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#24 AR-1248-4

R.T.: 5.537 min
Delta R.T.: -0.001 min
Response: 9780744
Conc: 25.74 ng/ml




#25 AR-1248-5

R.T.: 6.761 min
Delta R.T.: -0.065 min
Response: 2570146
Conc: 16.01 ng/ml



#25 AR-1248-5


R.T.: 5.892 min
Delta R.T.: -0.042 min
Response: 5892033
Conc: 15.83 ng/ml



#26 AR-1254-1


R.T.: 6.761 min
Delta R.T.: 0.000 min
Response: 2570146
Conc: 14.31 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#26 AR-1254-1

R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 5892033
Conc: 10.25 ng/ml




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 6.978 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.038 min
Delta R.T.: 0.000 min
Response: 7227196
Conc: 13.21 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 7.348 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#28 AR-1254-3

R.T.: 6.459 min
Delta R.T.: 0.014 min
Response: 24680092
Conc: 25.67 ng/ml




#30 AR-1254-5

R.T.: 8.053 min
Delta R.T.: 0.000 min
Response: 8003334
Conc: 34.51 ng/ml



#30 AR-1254-5


R.T.: 7.092 min
Delta R.T.: -0.002 min
Response: 24752861
Conc: 30.14 ng/ml



#31 AR-1260-1


R.T.: 7.511 min
Delta R.T.: 0.000 min
Response: 6689797
Conc: 32.16 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#31 AR-1260-1

R.T.: 6.576 min
Delta R.T.: 0.000 min
Response: 19615371
Conc: 33.91 ng/ml




#32 AR-1260-2

R.T.: 7.766 min
Delta R.T.: 0.000 min
Response: 7274801
Conc: 29.63 ng/ml



#32 AR-1260-2


R.T.: 6.761 min
Delta R.T.: -0.001 min
Response: 27243370
Conc: 34.51 ng/ml



#33 AR-1260-3


R.T.: 8.127 min
Delta R.T.: 0.000 min
Response: 5122861
Conc: 27.46 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#33 AR-1260-3

R.T.: 6.918 min
Delta R.T.: 0.000 min
Response: 20245209
Conc: 29.77 ng/ml




#34 AR-1260-4

R.T.: 8.367 min
Delta R.T.: 0.001 min
Response: 7159283
Conc: 33.99 ng/ml



#34 AR-1260-4


R.T.: 7.391 min
Delta R.T.: -0.002 min
Response: 15639764
Conc: 28.54 ng/ml



#35 AR-1260-5


R.T.: 8.705 min
Delta R.T.: 0.000 min
Response: 11673344
Conc: 27.78 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#35 AR-1260-5

R.T.: 7.631 min
Delta R.T.: -0.002 min
Response: 43564173
Conc: 29.56 ng/ml




#36 AR-1262-1

R.T.: 8.127 min
Delta R.T.: 0.000 min
Response: 5122861
Conc: 17.48 ng/ml



#36 AR-1262-1


R.T.: 7.184 min
Delta R.T.: -0.001 min
Response: 10498421
Conc: 25.75 ng/ml



#37 AR-1262-2


R.T.: 8.705 min
Delta R.T.: -0.001 min
Response: 11673344
Conc: 21.46 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#37 AR-1262-2

R.T.: 7.631 min
Delta R.T.: -0.002 min
Response: 43564173
Conc: 23.53 ng/ml




#38 AR-1262-3

R.T.: 9.058 min
Delta R.T.: 0.021 min
Response: 7658686
Conc: 22.36 ng/ml



#38 AR-1262-3


R.T.: 7.917 min
Delta R.T.: -0.001 min
Response: 8474136
Conc: 12.20 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 9.134 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#39 AR-1262-4

R.T.: 7.980 min
Delta R.T.: -0.004 min
Response: 30111292
Conc: 24.67 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 9.839 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.498 min
Delta R.T.: -0.002 min
Response: 9779889
Conc: 18.53 ng/ml



#41 AR-1268-1


R.T.: 9.058 min
Delta R.T.: 0.022 min
Response: 7658686
Conc: 13.97 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#41 AR-1268-1

R.T.: 7.917 min
Delta R.T.: -0.001 min
Response: 8474136
Conc: 4.51 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 9.136 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.980 min
Delta R.T.: -0.003 min
Response: 30111292
Conc: 18.04 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.381 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019




#43 AR-1268-3

R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 7094766
Conc: 5.05 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.839 min
Response: 0
Conc: N.D.



#44 AR-1268-4

R.T.: 8.498 min
Delta R.T.: -0.001 min
Response: 9779889
Conc: 18.07 ng/ml