

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR101819\  
Data File : PR042364.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 18 Oct 2019 18:29  
Operator : SM\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 19 00:15:31 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR101719.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Thu Oct 17 06:40:58 2019  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1   | RT#2  | Resp#1   | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-----------------|--------|-------|----------|---------|--------|--------|
| -----                       |                 |        |       |          |         |        |        |
| System Monitoring Compounds |                 |        |       |          |         |        |        |
| 1)                          | SA Tetrachlo... | 4.723  | 3.984 | 112.3E6  | 426.7E6 | 23.585 | 25.061 |
| 2)                          | SA Decachlor... | 10.650 | 9.096 | 74625361 | 231.9E6 | 21.344 | 22.848 |

Target Compounds

-----



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR101819\  
Data File : PR042364.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 18 Oct 2019 18:29  
Operator : SM\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 19 00:15:31 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR101719.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Thu Oct 17 06:40:58 2019  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.723 min  
Delta R.T.: 0.000 min  
Response: 112320454  
Conc: 23.58 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.984 min  
Delta R.T.: 0.000 min  
Response: 426680807  
Conc: 25.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.650 min  
Delta R.T.: 0.002 min  
Response: 74625361  
Conc: 21.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.096 min  
Delta R.T.: -0.002 min  
Response: 231923918  
Conc: 22.85 ng/ml