

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101920\
 Data File : PR047850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2020 08:23
 Operator : AJ\MA
 Sample : AR1221CCC400
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1221366

Manual Integrations
 APPROVED

Ankita
 10/20/2020 1:11:47 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 08:51:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 20 05:12:21 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.734	3.883	40582809	56216758	25.576m	19.920
2) SA Decachlor...	10.672	8.977	89823300	545.8E6	42.858	42.020
Target Compounds						
8) L2 AR-1221-1	4.942	4.096	8946387	11155041	417.936	342.434
9) L2 AR-1221-2	5.030	4.182	6411837	8514820	415.904	352.237
10) L2 AR-1221-3	5.109	4.260	21800620	27817067	426.056	377.067

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101920\
Data File : PR047850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2020 08:23
Operator : AJ\MA
Sample : AR1221CCC400
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1221366

Manual Integrations
APPROVED

Ankita
10/20/2020 1:11:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 08:51:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101920CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 20 05:12:21 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

