

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101920\
Data File : PR047829.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2020 03:04
Operator : AJ\MA
Sample : L4402-01DL 40X
Misc :
ALS Vial : 60 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AB4DL

Manual Integrations
APPROVED

Ankita
10/20/2020 1:11:19 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 08:05:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101920CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 20 05:12:21 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(21) AR-1248-1 (L5)

R.T.	Response	Conc
5.91	2302574	51.80
6.19	8076820	124.66
6.39	6567755	91.30
6.80	8312627	95.32
6.84	13076915	156.27

(21) AR-1248-1 #2 (L5)

R.T.	Response	Conc
5.00	9379981	201.94
5.21	8809345	151.30
5.26	6832317	95.06
5.43	7504561	77.56
5.83	27359801	175.15

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101920\
 Data File : PR047829.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2020 03:04
 Operator : AJ\MA
 Sample : L4402-01DL 40X
 Misc :
 ALS Vial : 60 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 C0AB4DL

Manual Integrations
 APPROVED

Ankita
 10/20/2020 1:11:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 08:05:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 20 05:12:21 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(21) AR-1248-1 #2 (L5)

R.T.	Response	Conc
5.91	2061005	46.36
6.19	8076820	124.66
6.39	6567755	91.30
6.80	8312627	95.32
6.84	13076915	156.27

(21) AR-1248-1 #2 (L5)

R.T.	Response	Conc
4.98	2139707	46.07
5.21	8809345	151.30
5.26	6832317	95.06
5.43	7504561	77.56
5.83	27359801	175.15

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101920\
 Data File : PR047829.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2020 03:04
 Operator : AJ\MA
 Sample : L4402-01DL 40X
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AB4DL

Manual Integrations
 APPROVED

Ankita
 10/20/2020 1:11:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 08:05:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 20 05:12:21 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

21) L5 AR-1248-1	5.912	4.982	2061005	2139707	46.365m	46.065m
22) L5 AR-1248-2	6.186	5.214	8076820	8809345	124.658	151.300
23) L5 AR-1248-3	6.394	5.257	6567755	6832317	91.302	95.058
24) L5 AR-1248-4	6.798	5.429	8312627	7504561	95.324	77.558
25) L5 AR-1248-5	6.838	5.825	13076915	27359801	156.269	175.150
31) L7 AR-1260-1	7.521	6.468	55293189	98910529	582.095	597.828
32) L7 AR-1260-2	7.777	6.656	84148265	378.1E6	706.507	709.055
33) L7 AR-1260-3	8.137	6.811	40481917	174.1E6	427.755	526.741
34) L7 AR-1260-4	8.376	7.287	54524485	165.9E6	501.129	405.685
35) L7 AR-1260-5	8.715	7.529	105.3E6	942.9E6	462.003	489.578

AJ
 10/24/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.