

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101924\
 Data File : PR068830.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2024 11:56
 Operator : AJ\MA
 Sample : P4413-09
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 KR-SS27

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 00:46:12 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 16 13:42:17 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.646	2.977	14853982	88218876	15.843	15.856
2) SA Decachlor...	9.537	8.264	9263256	53546381	16.054m	12.552m
Target Compounds						
31) L7 AR-1260-1	6.582	5.669	702911	5245263	18.405	22.981
32) L7 AR-1260-2	6.857	5.874	3016622	7550248	68.356	27.827 #
33) L7 AR-1260-3	7.253	6.034	969693	4530747	29.042m	18.191 #
34) L7 AR-1260-4	7.495	6.542	1024179	5349683	27.017m	24.866m
35) L7 AR-1260-5	7.834	6.807	2526981	15453569	36.771m	30.824m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR101924\
Data File : PR068830.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 11:56
Operator : AJ\MA
Sample : P4413-09
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
KR-SS27

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :Ankita Jodhani 10/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 00:46:12 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101624.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 16 13:42:17 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

