

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102022\
 Data File : PR057504.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2022 09:38
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR102022AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 10:38:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 21 06:41:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.695	2.925	134.8E6	77060924	52.215	50.886
2)	SA Decachlor...	9.642	8.193	117.3E6	104.7E6	48.371	50.932
Target Compounds							
21)	L5 AR-1248-1	4.925	4.044	31682113	14051521	490.567	475.970
22)	L5 AR-1248-2	5.219	4.292	41675078	17133626	504.225	476.844
23)	L5 AR-1248-3	5.433	4.334	49836100	17752315	511.359	469.342
24)	L5 AR-1248-4	5.868	4.509	56798932	23154710	513.288	470.795
25)	L5 AR-1248-5	5.907	4.921	54593013	24513156	517.739	468.494

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102022\
Data File : PR057504.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2022 09:38
Operator : AJ\MA
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR102022AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 10:38:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 21 06:41:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

