

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102022\
 Data File : PR057506.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2022 10:59
 Operator : AJ\MA
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR102022AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 11:13:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 21 06:41:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.697	2.924	137.3E6	80035184	53.191	52.850
2) SA Decachlor...	9.644	8.194	202.7E6	190.2E6	83.588	92.515
Target Compounds						
41) L9 AR-1268-1	8.216	7.045	187.1E6	127.4E6	453.133	464.403
42) L9 AR-1268-2	8.308	7.114	168.8E6	133.5E6	449.869	467.751
43) L9 AR-1268-3	8.529	7.339	138.4E6	114.5E6	439.860	475.608
44) L9 AR-1268-4	8.937	7.651	61377698	48514023	434.673	485.478
45) L9 AR-1268-5	9.327	7.948	440.8E6	444.7E6	426.253	477.915

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102022\
Data File : PR057506.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2022 10:59
Operator : AJ\MA
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR102022AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 11:13:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 21 06:41:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

