

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102120\
 Data File : PR047902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2020 01:44
 Operator : AJ\MA
 Sample : AR1254ICC400
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 09:59:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 21 09:58:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.731	3.882	42182483	63781917	20.000	20.000
2) SA Decachlor...	10.666	8.975	85914242	494.2E6	40.000	40.000
Target Compounds						
26) L6 AR-1254-1	6.769	5.782	38316073	66750731	400.000	400.000
27) L6 AR-1254-2	6.986	5.930	58816151	68290217	400.000	400.000
28) L6 AR-1254-3	7.356	6.337	62932455	149.5E6	400.000	400.000
29) L6 AR-1254-4	7.641	6.566	49036201	200.8E6	400.000	400.000
30) L6 AR-1254-5	8.062	6.986	53568664	260.5E6	400.000	400.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102120\
Data File : PR047902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2020 01:44
Operator : AJ\MA
Sample : AR1254ICC400
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 09:59:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 09:58:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 6.769 min
Delta R.T.: 0.000 min
Response: 38316073
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 5.782 min
Delta R.T.: 0.000 min
Response: 66750731
Conc: 400.00 ng/ml




#27 AR-1254-2

R.T.: 6.986 min
Delta R.T.: 0.000 min
Response: 58816151
Conc: 400.00 ng/ml



#27 AR-1254-2


R.T.: 5.930 min
Delta R.T.: 0.000 min
Response: 68290217
Conc: 400.00 ng/ml



#28 AR-1254-3


R.T.: 7.356 min
Delta R.T.: 0.000 min
Response: 62932455
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 6.337 min
Delta R.T.: 0.000 min
Response: 149481265
Conc: 400.00 ng/ml




#29 AR-1254-4

R.T.: 7.641 min
Delta R.T.: 0.000 min
Response: 49036201
Conc: 400.00 ng/ml



#29 AR-1254-4


R.T.: 6.566 min
Delta R.T.: 0.000 min
Response: 200775786
Conc: 400.00 ng/ml



#30 AR-1254-5

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 53568664
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :



#30 AR-1254-5

R.T.: 6.986 min
Delta R.T.: 0.000 min
Response: 260515700
Conc: 400.00 ng/ml