

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102220\
Data File : PR048038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2020 20:27
Operator : AJ\MA
Sample : L4451-19
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AJ7

Manual Integrations
APPROVED

Ankita
10/24/2020 9:13:43 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 04:14:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)

R.T.	Response	Conc
7.52	1554900	15.16
7.77	2203501	17.37
8.14	1277067	13.19
8.37	2269250	20.74
8.71	2832123	12.46

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
6.47	8712369	49.15
6.66	11316135	18.80
6.81	5260675	14.22
7.29	5585541	12.34
7.53	26382845	14.00

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102220\
Data File : PR048038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2020 20:27
Operator : AJ\MA
Sample : L4451-19
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AJ7

Manual Integrations
APPROVED

Ankita
10/24/2020 9:13:43 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 04:14:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
7.52	1554900	15.16
7.77	2203501	17.37
8.14	1277067	13.19
8.37	2269250	20.74
8.71	2832123	12.46

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
6.47	2748067	15.50
6.66	11316135	18.80
6.81	5260675	14.22
7.29	5585541	12.34
7.53	26382845	14.00

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102220\
 Data File : PR048038.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2020 20:27
 Operator : AJ\MA
 Sample : L4451-19
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AJ7

Manual Integrations
 APPROVED

Ankita
 10/24/2020 9:13:43 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 04:14:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 21 10:52:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.731	3.883	28652043	43893416	15.862	16.254
2) SA Decachlor...	10.665	8.981	63021596	363.1E6	31.289	32.458
Target Compounds						
31) L7 AR-1260-1	7.520	6.465	1554900	2748067	15.155	15.501m
32) L7 AR-1260-2	7.774	6.657	2203501	11316135	17.367	18.797
33) L7 AR-1260-3	8.136	6.812	1277067	5260675	13.191	14.222
34) L7 AR-1260-4	8.374	7.289	2269250	5585541	20.739	12.339 #
35) L7 AR-1260-5	8.715	7.531	2832123	26382845	12.463	14.004

AJ
 10/26/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.