

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR102418CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 24 03:59:33 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033090.D 1600 =PR033092.D 800 =PR033091.D
 200 =PR033089.D 100 =PR033088.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.622	1.597	1.410	1.684	1.790	1.620	E7	8.60
2) SA	Decachlorobiphenyl	1.221	1.084	1.055	1.220	1.434	1.203	E7	12.48
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	4.136	3.991	3.541	4.361	4.962	4.199	E5	12.43
22) L5	AR-1248-2	5.070	4.556	4.213	5.288	6.082	5.042	E5	14.25
23) L5	AR-1248-3	3.909	3.686	3.367	4.133	4.681	3.955	E5	12.50
24) L5	AR-1248-4	4.409	4.245	3.834	4.632	5.191	4.462	E5	11.24
25) L5	AR-1248-5	4.279	4.114	3.710	4.497	5.078	4.336	E5	11.65
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR033090.D 1600 =PR033092.D 800 =PR033091.D
 200 =PR033089.D 100 =PR033088.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	5.178	5.162	4.496	5.349	5.504	5.138	E7	7.49
2) SA	Decachlorobiphenyl	5.198	5.030	4.648	5.101	5.895	5.174	E7	8.77
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR102418CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 24 03:59:33 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033090.D 1600 =PR033092.D 800 =PR033091.D
 200 =PR033089.D 100 =PR033088.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	1.089	1.112	0.969	1.123	1.257	1.110 E6	9.26
22) L5	AR-1248-2	1.616	1.615	1.430	1.692	1.880	1.647 E6	9.84
23) L5	AR-1248-3	1.667	1.679	1.480	1.727	1.916	1.694 E6	9.20
24) L5	AR-1248-4	2.091	2.111	1.862	2.169	2.436	2.134 E6	9.63
25) L5	AR-1248-5	2.179	2.233	1.952	2.245	2.571	2.236 E6	9.91
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range