

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR102418CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 24 02:34:42 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033095.D 1600 =PR033097.D 800 =PR033096.D
 200 =PR033094.D 100 =PR033093.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.593	1.550	1.291	1.644	1.729	1.562 E7	10.57
2) SA	Decachlorobiphenyl	1.146	1.071	1.112	1.260	1.312	1.180 E7	8.64
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	4.606	4.099	3.858	4.951	5.205	4.544 E5	12.43
27) L6	AR-1254-2	7.898	7.057	6.578	8.524	9.303	7.872 E5	13.93
28) L6	AR-1254-3	0.898	0.834	0.771	0.951	1.042	0.899 E6	11.60
29) L6	AR-1254-4	6.606	6.060	5.621	6.898	7.519	6.541 E5	11.25
30) L6	AR-1254-5	7.424	6.919	6.434	7.737	8.259	7.355 E5	9.63
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

400 =PR033095.D 1600 =PR033097.D 800 =PR033096.D
 200 =PR033094.D 100 =PR033093.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.088	5.039	4.097	5.187	5.452	4.973 E7	10.36
2) SA	Decachlorobiphenyl	4.811	5.009	4.962	5.254	5.368	5.081 E7	4.45
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR102418CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 24 02:34:42 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033095.D 1600 =PR033097.D 800 =PR033096.D
 200 =PR033094.D 100 =PR033093.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	3.361	3.215	2.902	3.503	3.679	3.332 E6	8.87
27) L6	AR-1254-2	3.035	2.910	2.628	3.154	3.296	3.005 E6	8.46
28) L6	AR-1254-3	5.142	5.105	4.580	5.342	5.402	5.114 E6	6.34
29) L6	AR-1254-4	3.079	3.062	2.700	3.181	3.154	3.035 E6	6.39
30) L6	AR-1254-5	4.569	4.561	4.143	4.706	4.833	4.562 E6	5.70
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range