

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR102418CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 24 05:44:04 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033100.D 1600 =PR033102.D 800 =PR033101.D
 200 =PR033099.D 100 =PR033098.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.603	1.549	1.546	1.724	1.741	1.633	E7	5.77
2) SA	Decachlorobiphenyl	0.981	0.946	0.985	1.138	1.265	1.063	E7	12.71
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	0.850	0.829	0.842	0.954	1.008	0.897	E6	8.91
37) L8	AR-1262-2	1.457	1.485	1.485	1.619	1.724	1.554	E6	7.34
38) L8	AR-1262-3	0.975	0.960	0.983	1.109	1.201	1.046	E6	10.07
39) L8	AR-1262-4	7.372	7.189	7.378	8.372	9.075	7.877	E5	10.35
40) L8	AR-1262-5	4.633	4.522	4.678	5.354	5.970	5.032	E5	12.28
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR033100.D 1600 =PR033102.D 800 =PR033101.D
 200 =PR033099.D 100 =PR033098.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	5.128	4.999	4.944	5.489	5.657	5.243	E7	5.99
2) SA	Decachlorobiphenyl	4.427	4.627	4.627	5.018	5.661	4.872	E7	10.07
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR102418CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 24 05:44:04 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033100.D 1600 =PR033102.D 800 =PR033101.D
 200 =PR033099.D 100 =PR033098.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	5.276	5.437	5.356	5.787	5.994	5.570 E6	5.51
37) L8	AR-1262-2	0.887	0.962	0.929	0.967	1.003	0.950 E7	4.60
38) L8	AR-1262-3	3.171	3.296	3.267	3.532	3.758	3.405 E6	6.99
39) L8	AR-1262-4	6.047	6.454	6.309	6.669	7.137	6.523 E6	6.30
40) L8	AR-1262-5	2.465	2.537	2.548	2.797	3.122	2.694 E6	10.04
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range