Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102520\
Data File : PR048168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2020 16:38
Operator : AJ\MA
Sample : L4505-07
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0BG3

Manual Integrations APPROVED
mohammad
10/26/2020 5:08:49 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 01:28:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.727	3.880	35427306	52672865	19.613	19.505
2)	SA Decachlor...	10.660	8.978	59268605	318.2E6	29.425	28.439
	Target Compounds						
26)	L6 AR-1254-1	6.767	5.780	8600.6E6	12708.3E6	92722.692	79213.226
27)	L6 AR-1254-2	6.984	5.928	13563.6E6	15079.0E6	94136.881	92016.677
28)	L6 AR-1254-3	7.355	6.340	16293.1E6	38858.1E6	106259.190	108150.015
29)	L6 AR-1254-4	7.640	6.553	13618.8E6	45248.0E6	114836.086	94218.100m
30)	L6 AR-1254-5	8.061	7.003	15846.6E6	72235.6E6	124445.634	116110.238

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102520\
Data File : PR048168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2020 16:38
Operator : AJ\MA
Sample : L4505-07
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0BG3

Manual Integrations APPROVED
mohammad
10/26/2020 5:08:49 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 01:28:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm