Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102520\
Data File : PR048164.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2020 15:40
Operator : AJ\MA
Sample : L4505-03
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0BF3

Manual Integrations APPROVED
mohammad
10/26/2020 5:08:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 01:24:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.727	3.880	40436961	59722080	22.386	22.115
2)	SA Decachlor...	10.661	8.978	74271528	452.9E6	36.874	40.483
	Target Compounds						
26)	L6 AR-1254-1	6.766	5.779	5210.2E6	7421.9E6	56170.368	46261.699
27)	L6 AR-1254-2	6.983	5.927	8824.2E6	9021.8E6	61243.653	55053.732
28)	L6 AR-1254-3	7.354	6.336	10993.3E6	26858.8E6	71694.964	74753.543
29)	L6 AR-1254-4	7.640	6.556	9183.6E6	37562.0E6	77437.439	78213.863m
30)	L6 AR-1254-5	8.061	6.992	10790.6E6	56223.4E6	84739.721	90372.406

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102520\
Data File : PR048164.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2020 15:40
Operator : AJ\MA
Sample : L4505-03
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0BF3

Manual Integrations APPROVED
mohammad
10/26/2020 5:08:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 01:24:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm