

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102522\
 Data File : PR057696.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2022 00:36
 Operator : AJ\MA
 Sample : N5226-02
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 20221094-199-SITE-WATER

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 26 06:07:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 24 18:17:24 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.692	2.924	52024062	41491963	24.768	20.621
2) SA Decachlor...	9.630	8.177	42793600	46545185	19.355	19.714

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102522\
Data File : PR057696.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2022 00:36
Operator : AJ\MA
Sample : N5226-02
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
20221094-199-SITE-WATER

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 06:07:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 24 18:17:24 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.001 min
Response: 52024062
Conc: 24.77 ng/ml

Instrument :
ECD_R
ClientSampleId :
20221094-199-SITE-WATER




#1 Tetrachloro-m-xylene

R.T.: 2.924 min
Delta R.T.: -0.002 min
Response: 41491963
Conc: 20.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.630 min
Delta R.T.: -0.001 min
Response: 42793600
Conc: 19.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.177 min
Delta R.T.: -0.009 min
Response: 46545185
Conc: 19.71 ng/ml