

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102618\
 Data File : PR033207.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2018 02:44
 Operator : SM\SJ
 Sample : AR1254ICC800
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254ICC800

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 26 04:05:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR102618CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 26 03:58:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.759	3.740	806.9E6	2536.1E6	39.674	39.538
2)	SA Decachlor...	10.741	8.734	1260.6E6	3252.4E6	78.096	77.941
Target Compounds							
26)	L6 AR-1254-1	6.803	5.621	668.9E6	2904.0E6	776.274	778.212
27)	L6 AR-1254-2	7.022	5.765	1021.6E6	2550.0E6	768.331	779.504
28)	L6 AR-1254-3	7.393	6.168	1065.9E6	4126.7E6	761.300	763.272
29)	L6 AR-1254-4	7.680	6.394	764.1E6	2932.1E6	757.737	758.316
30)	L6 AR-1254-5	8.102	6.810	829.3E6	3122.1E6	752.270	732.967

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102618\
Data File : PR033207.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2018 02:44
Operator : SM\SJ
Sample : AR1254ICC800
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254ICC800

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 04:05:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR102618CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 26 03:58:23 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

