

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102620\
Data File : PR048203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2020 06:49
Operator : AJ\MA
Sample : L4505-05DL2 50X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0BF9DL2

Manual Integrations
APPROVED

Ankita
10/28/2020 3:41:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 07:41:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)
R.T. Response Conc
6.76 55580705 599.21
6.98 92939068 645.04
7.35 114430398 746.28
7.64 93445476 787.95
8.06 106228058 834.22
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.78 79933270 498.24
5.93 97662801 595.97
6.33 252696424 703.31
6.56 412645378 859.23
6.98 493405385 793.09

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102620\
Data File : PR048203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2020 06:49
Operator : AJ\MA
Sample : L4505-05DL2 50X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0BF9DL2

Manual Integrations
APPROVED

Ankita
10/28/2020 3:41:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 07:41:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.76 55580705 599.21
6.98 92939068 645.04
7.35 114430398 746.28
7.64 93445476 787.95
8.06 106228058 834.22

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.78 79933270 498.24
5.93 97662801 595.97
6.33 252696424 703.31
6.56 358073502 745.60
6.98 493405385 793.09

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102620\
 Data File : PR048203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2020 06:49
 Operator : AJ\MA
 Sample : L4505-05DL2 50X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0BF9DL2

Manual Integrations
 APPROVED

Ankita
 10/28/2020 3:41:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 07:41:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 21 10:52:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

26)	L6	AR-1254-1	6.764	5.778	55580705	79933270	599.212	498.238
27)	L6	AR-1254-2	6.981	5.926	92939068	97662801	645.037	595.970
28)	L6	AR-1254-3	7.353	6.333	114.4E6	252.7E6	746.283	703.306
29)	L6	AR-1254-4	7.638	6.562	93445476	358.1E6	787.949	745.602m
30)	L6	AR-1254-5	8.058	6.984	106.2E6	493.4E6	834.222	793.091

S. AJ
 10/30/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.