

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102620\
Data File : PR048238.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2020 15:15
Operator : AJ\MA
Sample : L4505-13DL2 50X
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0BH2DL2

Manual Integrations
APPROVED

Ankita
10/28/2020 3:59:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 02:05:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)
R.T. Response Conc
6.76 81323775 876.75
6.98 138444443 960.86
7.35 172613764 1125.74
7.64 137273902 1157.52
8.06 154022220 1209.56
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.78 123101031 767.31
5.93 147687295 901.24
6.33 374932538 1043.51
6.56 617380944 1285.55
6.99 726333270 1167.49

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102620\
Data File : PR048238.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2020 15:15
Operator : AJ\MA
Sample : L4505-13DL2 50X
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
C0BH2DL2

Manual Integrations
APPROVED

Ankita
10/28/2020 3:59:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 02:05:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.76 81323775 876.75
6.98 138444443 960.86
7.35 172613764 1125.74
7.64 137273902 1157.52
8.06 154022220 1209.56

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.78 123101031 767.31
5.93 147687295 901.24
6.33 374932538 1043.51
6.56 545034186 1134.90
6.99 726333270 1167.49

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102620\
 Data File : PR048238.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2020 15:15
 Operator : AJ\MA
 Sample : L4505-13DL2 50X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0BH2DL2

Manual Integrations
 APPROVED

Ankita
 10/28/2020 3:59:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 02:05:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 21 10:52:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

26)	L6	AR-1254-1	6.764	5.778	81323775	123.1E6	876.747	767.310
27)	L6	AR-1254-2	6.981	5.926	138.4E6	147.7E6	960.864	901.236
28)	L6	AR-1254-3	7.352	6.334	172.6E6	374.9E6	1125.739	1043.514
29)	L6	AR-1254-4	7.638	6.563	137.3E6	545.0E6	1157.518	1134.902m
30)	L6	AR-1254-5	8.058	6.985	154.0E6	726.3E6	1209.556	1167.495

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

AJ
 10/30/20