

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR102620\  
Data File : PR048244.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 27 Oct 2020 16:56  
Operator : AJ\MA  
Sample : L4505-17DL2 500X  
Misc :  
ALS Vial : 13 Sample Multiplier: 1


Instrument :  
ECD\_R  
ClientSampleId :  
C0BH4DL2

Manual Integrations  
APPROVED

Ankita  
10/28/2020 3:59:48 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 28 02:10:24 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR102120CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Oct 21 10:52:16 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)  
R.T. Response Conc  
6.76 57705619 622.12  
6.98 101602287 705.16  
7.35 126502432 825.01  
7.64 105697825 891.26  
8.06 113291189 889.69

(26) AR-1254-1 #2 (L6)  
R.T. Response Conc  
5.78 83390925 519.79  
5.93 109073361 665.60  
6.33 275139418 765.77  
6.56 463933921 966.03  
6.98 522546116 839.93

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR102620\  
Data File : PR048244.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 27 Oct 2020 16:56  
Operator : AJ\MA  
Sample : L4505-17DL2 500X  
Misc :  
ALS Vial : 13 Sample Multiplier: 1


Instrument :  
ECD\_R  
ClientSampleId :  
C0BH4DL2

Manual Integrations  
APPROVED

Ankita  
10/28/2020 3:59:48 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 28 02:10:24 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR102120CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Oct 21 10:52:16 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)  
R.T. Response Conc  
6.76 57705619 622.12  
6.98 101602287 705.16  
7.35 126502432 825.01  
7.64 105697825 891.26  
8.06 113291189 889.69

(26) AR-1254-1 #2 (L6)  
R.T. Response Conc  
5.78 83390925 519.79  
5.93 109073361 665.60  
6.33 275139418 765.77  
6.56 408139532 849.85  
6.98 522546116 839.93

## Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR102620\  
Data File : PR048244.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 27 Oct 2020 16:56  
Operator : AJ\MA  
Sample : L4505-17DL2 500X  
Misc :  
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 28 02:10:24 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR102120CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Oct 21 10:52:16 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound | RT#1 | RT#2 | Resp#1 | Resp#2 | ng/ml | ng/ml |
|----------|------|------|--------|--------|-------|-------|
|----------|------|------|--------|--------|-------|-------|

## System Monitoring Compounds

## Target Compounds

|     |    |           |       |       |          |          |         |          |
|-----|----|-----------|-------|-------|----------|----------|---------|----------|
| 26) | L6 | AR-1254-1 | 6.764 | 5.778 | 57705619 | 83390925 | 622.121 | 519.790  |
| 27) | L6 | AR-1254-2 | 6.981 | 5.926 | 101.6E6  | 109.1E6  | 705.163 | 665.601  |
| 28) | L6 | AR-1254-3 | 7.352 | 6.333 | 126.5E6  | 275.1E6  | 825.014 | 765.769  |
| 29) | L6 | AR-1254-4 | 7.637 | 6.563 | 105.7E6  | 408.1E6  | 891.263 | 849.852m |
| 30) | L6 | AR-1254-5 | 8.059 | 6.984 | 113.3E6  | 522.5E6  | 889.690 | 839.931  |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :  
ECD\_R  
ClientSampleId :  
C0BH4DL2

Manual Integrations  
APPROVED

Ankita  
10/28/2020 3:59:48 PM

AJ  
10/30/20