

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102718\
 Data File : PR033288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2018 23:50
 Operator : SM\SJ
 Sample : MDL-AR1262-S-1
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1262-S-1

Manual Integrations
 APPROVED

Sohil
 10/30/2018 11:25:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 00:43:43 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR102718CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 27 00:40:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.759	3.740	403.0E6	1204.1E6	19.874	19.050
2) SA Decachlor...	10.744	8.732	634.3E6	1572.0E6	39.305	38.322
Target Compounds						
36) L8 AR-1262-1	8.177	6.843	37863108	93323057	30.973	22.135 #
37) L8 AR-1262-2	8.764	7.343	58581519	192.8E6	28.590	26.083m
38) L8 AR-1262-3	9.103	7.626	39299470	76856020	28.572	29.289
39) L8 AR-1262-4	9.198	7.690	29815165	142.9E6	28.510	28.164
40) L8 AR-1262-5	9.919	8.183	19604358	60228459	27.743m	27.410

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102718\
Data File : PR033288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2018 23:50
Operator : SM\SJ
Sample : MDL-AR1262-S-1
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1262-S-1

Manual Integrations
APPROVED

Sohil
10/30/2018 11:25:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 00:43:43 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR102718CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 27 00:40:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

