

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102718\
 Data File : PR033301.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2018 02:57
 Operator : SM\SJ
 Sample : MDL-AR1262-W-6
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1262-W-6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 03:58:47 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR102718CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 27 00:40:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.760	3.742	410.4E6	1228.3E6	20.236	19.433
2) SA Decachlor...	10.745	8.733	890.8E6	2235.3E6	55.196	54.491
Target Compounds						
36) L8 AR-1262-1	8.178	6.844	51806402	129.7E6	42.379	30.755 #
37) L8 AR-1262-2	8.764	7.343	83247585	280.5E6	40.629	37.953
38) L8 AR-1262-3	9.102	7.625	58149968	109.8E6	42.276	41.832
39) L8 AR-1262-4	9.198	7.689	43665401	199.1E6	41.754	39.246
40) L8 AR-1262-5	9.918	8.182	29074099	84118533	41.144	38.282

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102718\
Data File : PR033301.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2018 02:57
Operator : SM\SJ
Sample : MDL-AR1262-W-6
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1262-W-6

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 03:58:47 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR102718CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 27 00:40:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

