

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102820\
 Data File : PR048376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2020 02:14
 Operator : AJ\MA
 Sample : AIBLK96
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 05:56:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 21 10:52:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.726	3.878	45298380	60291646	25.077	22.326
2) SA Decachlor...	10.655	8.979	77512105	426.0E6	38.483	38.074
Target Compounds						
25) L5 AR-1248-5	0.000	5.878f	0	27972290	N.D.	185.628 #
27) L6 AR-1254-2	0.000	5.912	0	15009401	N.D.	91.592 #
29) L6 AR-1254-4	0.000	6.539	0	3056560	N.D.	6.365 #
34) L7 AR-1260-4	8.388	0.000	30251471	0	276.472	N.D. #
43) L9 AR-1268-3	0.000	8.089	0	11366905	N.D.	6.088 #
45) L9 AR-1268-5	0.000	8.707	0	9249242	N.D.	1.713 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102820\
Data File : PR048376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2020 02:14
Operator : AJ\MA
Sample : AIBLK96
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 05:56:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.835 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#25 AR-1248-5

R.T.: 5.878 min
Delta R.T.: 0.054 min
Response: 27972290
Conc: 185.63 ng/ml




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T.: 6.986 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 5.912 min
Delta R.T.: -0.017 min
Response: 15009401
Conc: 91.59 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 7.641 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#29 AR-1254-4

R.T.: 6.539 min
Delta R.T.: -0.026 min
Response: 3056560
Conc: 6.36 ng/ml




#34 AR-1260-4

R.T.: 8.388 min
Delta R.T.: 0.008 min
Response: 30251471
Conc: 276.47 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 7.290 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T. : 9.390 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#43 AR-1268-3

R.T.: 8.089 min
Delta R.T.: -0.002 min
Response: 11366905
Conc: 6.09 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 10.296 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.707 min
Delta R.T.: 0.005 min
Response: 9249242
Conc: 1.71 ng/ml