

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102820\
Data File : PR048379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2020 02:57
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1254389

Manual Integrations
APPROVED

Ankita
10/29/2020 12:05:35 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 05:58:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
6.76 36868932 397.48
6.98 54963256 381.47
7.35 55713562 363.35
7.64 42569644 358.95
8.06 44764495 351.54
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.78 58107811 362.20
5.92 57151450 348.76
6.33 121977654 339.49
6.56 186426001 388.19
6.98 205723549 330.68

QEdit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102820\
Data File : PR048379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2020 02:57
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1254389

Manual Integrations
APPROVED

Ankita
10/29/2020 12:05:35 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 05:58:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 #2 (L6)	R.T.	Response	Conc
	6.76	36868932	397.48
	6.98	54963256	381.47
	7.35	55713562	363.35
	7.64	42569644	358.95
	8.06	44764495	351.54

(26) AR-1254-1 #2 (L6)	R.T.	Response	Conc
	5.78	58107811	362.20
	5.92	57151450	348.76
	6.33	121977654	339.49
	6.56	159987085	333.13
	6.98	205723549	330.68

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102820\
 Data File : PR048379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2020 02:57
 Operator : AJ\MA
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1254389

Manual Integrations
 APPROVED

Ankita
 10/29/2020 12:05:35 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 05:58:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 21 10:52:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.726	3.878	47718636	63786821	26.417	23.620
2) SA Decachlor...	10.656	8.980	82156245	456.6E6	40.789	40.815
Target Compounds						
26) L6 AR-1254-1	6.763	5.777	36868932	58107811	397.482	362.196
27) L6 AR-1254-2	6.980	5.925	54963256	57151450	381.469	348.757
28) L6 AR-1254-3	7.350	6.332	55713562	122.0E6	363.348	339.489
29) L6 AR-1254-4	7.636	6.562	42569644	160.0E6	358.955	333.135m
30) L6 AR-1254-5	8.057	6.984	44764495	205.7E6	351.541	330.676

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

AJ
 11/09/20