

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102822\
 Data File : PR057737.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2022 18:18
 Operator : AJ\MA
 Sample : AR1242ICC800
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 04:48:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102822CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.692	2.923	102.4E6	82810547	38.410	40.236
2) SA Decachlor...	9.640	8.178	149.4E6	192.9E6	75.368	77.687
Target Compounds						
16) L4 AR-1242-1	4.921	4.039	50008007	37451300	738.327	774.485
17) L4 AR-1242-2	4.943	4.057	68420160	58475365	744.850	781.720
18) L4 AR-1242-3	5.008	4.235	43226145	32585455	738.134	776.761
19) L4 AR-1242-4	5.111	4.328	36521204	25715415	739.087	755.978
20) L4 AR-1242-5	5.903	4.872	40312380	35321698	670.291	768.459



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102822\
Data File : PR057737.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2022 18:18
Operator : AJ\MA
Sample : AR1242ICC800
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 04:48:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102822CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#16 AR-1242-1


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 50008007
Conc: 738.33 ng/ml

Instrument :
ECD_R
ClientSampleId :




#16 AR-1242-1

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 37451300
Conc: 774.49 ng/ml




#17 AR-1242-2

R.T.: 4.943 min
Delta R.T.: 0.000 min
Response: 68420160
Conc: 744.85 ng/ml



#17 AR-1242-2


R.T.: 4.057 min
Delta R.T.: 0.000 min
Response: 58475365
Conc: 781.72 ng/ml



#18 AR-1242-3


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 43226145
Conc: 738.13 ng/ml

Instrument :
ECD_R
ClientSampleId :




#18 AR-1242-3

R.T.: 4.235 min
Delta R.T.: 0.000 min
Response: 32585455
Conc: 776.76 ng/ml




#19 AR-1242-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 36521204
Conc: 739.09 ng/ml



#19 AR-1242-4


R.T.: 4.328 min
Delta R.T.: 0.000 min
Response: 25715415
Conc: 755.98 ng/ml



#20 AR-1242-5

R.T.: 5.903 min
Delta R.T.: 0.000 min
Response: 40312380
Conc: 670.29 ng/ml

Instrument :
ECD_R
ClientSampleId :



#20 AR-1242-5

R.T.: 4.872 min
Delta R.T.: 0.000 min
Response: 35321698
Conc: 768.46 ng/ml