

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102822\
 Data File : PR057742.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2022 19:36
 Operator : AJ\MA
 Sample : AR1248ICC800
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 05:01:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102822CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 04:59:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.692	2.923	99576100	82911514	38.037	39.070
2) SA Decachlor...	9.640	8.178	149.6E6	193.0E6	75.232	78.586
Target Compounds						
21) L5 AR-1248-1	4.921	4.039	37790737	28012475	736.367	765.055
22) L5 AR-1248-2	5.215	4.286	49009614	34949700	731.219	759.456
23) L5 AR-1248-3	5.428	4.327	59391840	38101395	742.822	760.362
24) L5 AR-1248-4	5.863	4.502	69393599	49106763	751.687	762.908
25) L5 AR-1248-5	5.902	4.913	67595054	54359591	759.366	778.372



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102822\
Data File : PR057742.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2022 19:36
Operator : AJ\MA
Sample : AR1248ICC800
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 05:01:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102822CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 04:59:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#21 AR-1248-1


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 37790737
Conc: 736.37 ng/ml

Instrument :
ECD_R
ClientSampleId :




#21 AR-1248-1

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 28012475
Conc: 765.06 ng/ml




#22 AR-1248-2

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 49009614
Conc: 731.22 ng/ml



#22 AR-1248-2


R.T.: 4.286 min
Delta R.T.: 0.000 min
Response: 34949700
Conc: 759.46 ng/ml



#23 AR-1248-3


R.T.: 5.428 min
Delta R.T.: 0.000 min
Response: 59391840
Conc: 742.82 ng/ml

Instrument :
ECD_R
ClientSampleId :




#23 AR-1248-3

R.T.: 4.327 min
Delta R.T.: 0.000 min
Response: 38101395
Conc: 760.36 ng/ml




#24 AR-1248-4

R.T.: 5.863 min
Delta R.T.: 0.000 min
Response: 69393599
Conc: 751.69 ng/ml



#24 AR-1248-4


R.T.: 4.502 min
Delta R.T.: 0.000 min
Response: 49106763
Conc: 762.91 ng/ml



#25 AR-1248-5

R.T.: 5.902 min
Delta R.T.: -0.001 min
Response: 67595054
Conc: 759.37 ng/ml

Instrument :
ECD_R
ClientSampleId :



#25 AR-1248-5

R.T.: 4.913 min
Delta R.T.: 0.000 min
Response: 54359591
Conc: 778.37 ng/ml