

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102920\
 Data File : PR048386.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2020 10:31
 Operator : AJ\MA
 Sample : L4506-02DL2 500X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0BF6DL2

Manual Integrations
 APPROVED

Ankita
 11/2/2020 11:03:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 04:10:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 21 10:52:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
26)	L6	AR-1254-1	6.764	5.777	72313077	86502363	779.603 539.184 #
27)	L6	AR-1254-2	6.982	5.926	118.3E6	124.7E6	821.023 760.939
28)	L6	AR-1254-3	7.353	6.334	140.4E6	296.6E6	915.819 825.389
29)	L6	AR-1254-4	7.639	6.563	121.1E6	495.0E6	1021.304 1030.797m
30)	L6	AR-1254-5	8.061	6.986	115.0E6	535.8E6	903.230 861.262

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102920\
Data File : PR048386.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2020 10:31
Operator : AJ\MA
Sample : L4506-02DL2 500X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
C0BF6DL2

Manual Integrations
APPROVED

Ankita
11/2/2020 11:03:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 04:10:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 21 10:52:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

