

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103023\
Data File : PR064280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2023 12:08
Operator : AJ\MA
Sample : PB156736BS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
ALCS736

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/01/2023
Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 21:21:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 13 05:04:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(1) Tetrachloro-m-xylene (SA)

3.736min 19.377 ng/ml

response 138777342

(1) Tetrachloro-m-xylene #2 (SA)

3.021min 17.912 ng/ml

response 73932840

(+) = Expected Retention Time

PR101223CLP.M Wed Nov 01 06:34:26 2023

Page: 1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103023\
Data File : PR064280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2023 12:08
Operator : AJ\MA
Sample : PB156736BS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
ALCS736

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/01/2023
Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 21:21:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 13 05:04:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(1) Tetrachloro-m-xylene (SA)

3.736min 19.377 ng/ml

response 138777342

(1) Tetrachloro-m-xylene #2 (SA)

3.021min 18.291 ng/ml m

response 75494261

(+) = Expected Retention Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103023\
 Data File : PR064280.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2023 12:08
 Operator : AJ\MA
 Sample : PB156736BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :

ALCS736


Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/01/2023

Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 21:21:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 13 05:04:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.71	30468880	102.52
7.00	35447379	102.50
7.39	22936886	89.27
7.63	32372774	114.55
7.98	46831877	86.75

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.76	22782639	101.47
5.97	26537326	100.63
6.14	24472990	96.78
6.65	16497274	83.62
6.92	36614884	86.23

QEdit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103023\
 Data File : PR064280.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2023 12:08
 Operator : AJ\MA
 Sample : PB156736BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :

ALCS736


Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/01/2023

Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 21:21:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 13 05:04:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.71	30468880	102.52
7.00	35447379	102.50
7.39	22936886	89.27
7.63	26958445	95.39
7.98	46831877	86.75

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.76	22782639	101.47
5.97	26537326	100.63
6.14	24472990	96.78
6.65	16497274	83.62
6.92	36614884	86.23

QEdit

(+) = Expected Retention Time