

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103023\
 Data File : PR064283.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2023 12:52
 Operator : AJ\MA
 Sample : 05044-03MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BHEE3MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/01/2023
 Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 21:24:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 13 05:04:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.735	3.021	164.1E6	85353724	22.911	20.680m
2)	SA Decachlor...	9.755	8.413	106.8E6	68598035	22.863	22.599
Target Compounds							
3)	L1 AR-1016-1	4.977	4.170	102.2E6	61488322	475.076	422.981
4)	L1 AR-1016-2	5.000	4.189	151.0E6	85329126	460.779	426.848
5)	L1 AR-1016-3	5.065	4.373	95508393	46490911	461.687	421.720
6)	L1 AR-1016-4	5.168	4.424	76274903	37808089	468.276	436.896
7)	L1 AR-1016-5	5.487	4.648	75182549	47888234	462.505	419.164
31)	L7 AR-1260-1	6.711	5.764	131.6E6	96373772	442.725	429.224
32)	L7 AR-1260-2	6.996	5.972	153.2E6	113.1E6	442.858	428.789
33)	L7 AR-1260-3	7.388	6.135	100.2E6	106.8E6	390.057	422.293
34)	L7 AR-1260-4	7.633	6.651	123.1E6	63742077	435.478	323.102 #
35)	L7 AR-1260-5	7.974	6.920	218.0E6	170.9E6	403.807	402.503

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103023\
Data File : PR064283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2023 12:52
Operator : AJ\MA
Sample : 05044-03MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :

BHEE3MSD

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/01/2023

Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 21:24:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 13 05:04:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

