

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR110118.M
 Title : GC EXTRACTABLES
 Last Update : Thu Nov 01 18:15:24 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR033510.D 1000 =PR033508.D 750 =PR033509.D
 250 =PR033511.D 50 =PR033512.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.768	1.767	1.758	1.778	2.054	1.825	E7	7.01
2) SA	Decachlorobiphenyl	1.773	1.678	1.722	1.819	2.195	1.838	E7	11.26
3) L1	AR-1016-1	7.802	7.730	7.770	7.639	9.322	8.052	E5	8.85
4) L1	AR-1016-2	1.154	1.145	1.148	1.154	1.384	1.197	E6	8.74
5) L1	AR-1016-3	6.911	6.782	6.854	6.854	8.330	7.146	E5	9.28
6) L1	AR-1016-4	5.753	5.721	5.752	5.703	6.689	5.923	E5	7.23
7) L1	AR-1016-5	5.969	5.843	5.936	5.859	6.836	6.089	E5	6.92
8) L2	AR-1221-1	1.868					1.868	E5	0.00
9) L2	AR-1221-2	1.362					1.362	E5	0.00
10) L2	AR-1221-3	4.374					4.374	E5	0.00
11) L3	AR-1232-1	3.340					3.340	E5	0.00
12) L3	AR-1232-2	1.969					1.969	E5	0.00
13) L3	AR-1232-3	3.883					3.883	E5	0.00
14) L3	AR-1232-4	1.913					1.913	E5	0.00
15) L3	AR-1232-5	1.433					1.433	E5	0.00
16) L4	AR-1242-1	5.250					5.250	E5	0.00
17) L4	AR-1242-2	7.830					7.830	E5	0.00
18) L4	AR-1242-3	4.621					4.621	E5	0.00
19) L4	AR-1242-4	3.746					3.746	E5	0.00
20) L4	AR-1242-5	5.821					5.821	E5	0.00
21) L5	AR-1248-1	4.073					4.073	E5	0.00
22) L5	AR-1248-2	5.836					5.836	E5	0.00
23) L5	AR-1248-3	6.673					6.673	E5	0.00
24) L5	AR-1248-4	8.275					8.275	E5	0.00
25) L5	AR-1248-5	8.103					8.103	E5	0.00
26) L6	AR-1254-1	9.898					9.898	E5	0.00
27) L6	AR-1254-2	1.486					1.486	E6	0.00
28) L6	AR-1254-3	1.458					1.458	E6	0.00
29) L6	AR-1254-4	9.064					9.064	E5	0.00
30) L6	AR-1254-5	1.391					1.391	E6	0.00
31) L7	AR-1260-1	1.183	1.012	1.181	1.173	1.365	1.183	E6	10.59
32) L7	AR-1260-2	1.410	1.376	1.401	1.386	1.623	1.439	E6	7.21
33) L7	AR-1260-3	1.008	0.979	0.997	0.986	1.183	1.031	E6	8.33
34) L7	AR-1260-4	1.171	1.146	1.167	1.155	1.350	1.198	E6	7.13
35) L7	AR-1260-5	2.357	2.330	2.356	2.310	2.635	2.398	E6	5.60
36) L8	AR-1262-1	1.237					1.237	E6	0.00
37) L8	AR-1262-2	2.242					2.242	E6	0.00
38) L8	AR-1262-3	1.478					1.478	E6	0.00
39) L8	AR-1262-4	1.128					1.128	E6	0.00
40) L8	AR-1262-5	7.697					7.697	E5	0.00
41) L9	AR-1268-1	2.627					2.627	E6	0.00
42) L9	AR-1268-2	2.429					2.429	E6	0.00
43) L9	AR-1268-3	2.078					2.078	E6	0.00
44) L9	AR-1268-4	8.538					8.538	E5	0.00
45) L9	AR-1268-5	6.494					6.494	E6	0.00

Signal #2 Calibration Files

500 =PR033510.D 1000 =PR033508.D 750 =PR033509.D
 250 =PR033511.D 50 =PR033512.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	6.492	6.488	6.465	6.460	7.133	6.608	E7	4.45
2) SA	Decachlorobiphenyl	5.159	5.182	5.149	5.064	5.713	5.253	E7	4.97
3) L1	AR-1016-1	2.517	2.481	2.519	2.169	2.828	2.503	E6	9.33
4) L1	AR-1016-2	3.858	3.837	3.853	3.831	4.275	3.931	E6	4.90
5) L1	AR-1016-3	1.933	1.935	1.939	1.853	2.159	1.964	E6	5.84

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR110118.M
 Title : GC EXTRACTABLES
 Last Update : Thu Nov 01 18:15:24 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR033510.D 1000 =PR033508.D 750 =PR033509.D
 250 =PR033511.D 50 =PR033512.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.479	1.452	1.473	1.428	1.964	1.559	E6	14.58
7)	L1 AR-1016-5	2.026	2.010	2.023	2.008	2.313	2.076	E6	6.40
8)	L2 AR-1221-1	5.886					5.886	E5	0.00
9)	L2 AR-1221-2	4.419					4.419	E5	0.00
10)	L2 AR-1221-3	1.542					1.542	E6	0.00
11)	L3 AR-1232-1	1.316					1.316	E6	0.00
12)	L3 AR-1232-2	1.345					1.345	E6	0.00
13)	L3 AR-1232-3	6.340					6.340	E5	0.00
14)	L3 AR-1232-4	5.361					5.361	E5	0.00
15)	L3 AR-1232-5	7.532					7.532	E5	0.00
16)	L4 AR-1242-1	1.745					1.745	E6	0.00
17)	L4 AR-1242-2	2.574					2.574	E6	0.00
18)	L4 AR-1242-3	1.278					1.278	E6	0.00
19)	L4 AR-1242-4	1.193					1.193	E6	0.00
20)	L4 AR-1242-5	1.718					1.718	E6	0.00
21)	L5 AR-1248-1	1.275					1.275	E6	0.00
22)	L5 AR-1248-2	1.706					1.706	E6	0.00
23)	L5 AR-1248-3	1.736					1.736	E6	0.00
24)	L5 AR-1248-4	2.202					2.202	E6	0.00
25)	L5 AR-1248-5	2.260					2.260	E6	0.00
26)	L6 AR-1254-1	4.108					4.108	E6	0.00
27)	L6 AR-1254-2	3.596					3.596	E6	0.00
28)	L6 AR-1254-3	6.058					6.058	E6	0.00
29)	L6 AR-1254-4	3.815					3.815	E6	0.00
30)	L6 AR-1254-5	5.241					5.241	E6	0.00
31)	L7 AR-1260-1	3.931	3.861	3.932	3.806	4.469	4.000	E6	6.69
32)	L7 AR-1260-2	5.038	4.947	5.042	4.749	5.309	5.017	E6	4.03
33)	L7 AR-1260-3	4.566	4.494	4.583	4.402	4.863	4.582	E6	3.77
34)	L7 AR-1260-4	3.640	3.561	3.625	3.558	4.104	3.698	E6	6.22
35)	L7 AR-1260-5	0.943	0.951	0.952	0.910	1.012	0.953	E7	3.87
36)	L8 AR-1262-1	4.399					4.399	E6	0.00
37)	L8 AR-1262-2	9.030					9.030	E6	0.00
38)	L8 AR-1262-3	3.488					3.488	E6	0.00
39)	L8 AR-1262-4	6.275					6.275	E6	0.00
40)	L8 AR-1262-5	2.549					2.549	E6	0.00
41)	L9 AR-1268-1	1.078					1.078	E7	0.00
42)	L9 AR-1268-2	9.684					9.684	E6	0.00
43)	L9 AR-1268-3	8.323					8.323	E6	0.00
44)	L9 AR-1268-4	3.135					3.135	E6	0.00
45)	L9 AR-1268-5	2.386					2.386	E7	0.00

(#) = Out of Range