Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110122\
Data File : PR057823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2022 02:39
Operator : AJ\MA
Sample : AR1268ICC800
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12684244

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 10:54:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110122CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 01 10:53:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.694	2.921	86250748	72086448	36.930	38.798
2)	SA Decachlor...	9.646	8.171	286.7E6	412.6E6	74.573	77.660
	Target Compounds						
41)	L9 AR-1268-1	8.216	7.028	257.0E6	286.8E6	754.917	784.393
42)	L9 AR-1268-2	8.307	7.096	233.0E6	319.2E6	753.856	796.179
43)	L9 AR-1268-3	8.529	7.319	195.9E6	279.1E6	751.822	786.638
44)	L9 AR-1268-4	8.937	7.631	88822097	112.7E6	740.885	771.351
45)	L9 AR-1268-5	9.329	7.927	666.3E6	1080.7E6	773.489	797.462

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110122\
Data File : PR057823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2022 02:39
Operator : AJ\MA
Sample : AR1268ICC800
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12684244

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 10:54:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110122CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 01 10:53:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm