

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110219\
 Data File : PR042787.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2019 22:54
 Operator : SM\AJ
 Sample : AIBLK57
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AIBLK03

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 01:10:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110219CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 02 01:05:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.710	3.970	82240335	286.8E6	14.654	14.843
2)	SA Decachlor...	10.622	9.070	124.6E6	396.4E6	34.597	35.248
Target Compounds							
7)	L1 AR-1016-5	0.000	5.504	0	2858349	N.D.	9.014 #
9)	L2 AR-1221-2	0.000	4.277	0	4068993	N.D.	30.331 #
11)	L3 AR-1232-1	0.000	4.277f	0	4068993	N.D.	10.309 #
15)	L3 AR-1232-5	0.000	5.504	0	2858349	N.D.	20.975 #
20)	L4 AR-1242-5	0.000	5.900	0	31880161	N.D.	107.284 #
24)	L5 AR-1248-4	0.000	5.504	0	2858349	N.D.	6.862 #
25)	L5 AR-1248-5	0.000	5.920	0	3437102	N.D.	7.802 #
26)	L6 AR-1254-1	0.000	5.900	0	31880161	N.D.	51.061 #
27)	L6 AR-1254-2	0.000	6.021	0	14065720	N.D.	25.041 #
28)	L6 AR-1254-3	0.000	6.433	0	4098733	N.D.	4.197 #
29)	L6 AR-1254-4	0.000	6.598f	0	8696535	N.D.	12.170 #
31)	L7 AR-1260-1	0.000	6.598f	0	8696535	N.D.	13.278 #
32)	L7 AR-1260-2	0.000	6.741	0	1225547	N.D.	1.342 #
43)	L9 AR-1268-3	0.000	8.175	0	2710921	N.D.	1.635 #
45)	L9 AR-1268-5	0.000	8.793	0	3209700	N.D.	0.681 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110219\
Data File : PR042787.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2019 22:54
Operator : SM\AJ
Sample : AIBLK57
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AIBLK03

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 01:10:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 02 01:05:46 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#7 AR-1016-5


R.T.: 0.000 min
Exp R.T. : 6.367 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03




#7 AR-1016-5

R.T.: 5.504 min
Delta R.T.: -0.016 min
Response: 2858349
Conc: 9.01 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 5.006 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.277 min
Delta R.T.: 0.007 min
Response: 4068993
Conc: 30.33 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 5.085 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03




#11 AR-1232-1

R.T.: 4.277 min
Delta R.T.: -0.070 min
Response: 4068993
Conc: 10.31 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 6.160 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.504 min
Delta R.T.: -0.016 min
Response: 2858349
Conc: 20.97 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 6.811 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03




#20 AR-1242-5

R.T.: 5.900 min
Delta R.T.: 0.028 min
Response: 31880161
Conc: 107.28 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T.: 6.772 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.504 min
Delta R.T.: -0.016 min
Response: 2858349
Conc: 6.86 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 6.811 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03




#25 AR-1248-5

R.T.: 5.920 min
Delta R.T.: 0.005 min
Response: 3437102
Conc: 7.80 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.745 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.900 min
Delta R.T.: 0.027 min
Response: 31880161
Conc: 51.06 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T. : 6.962 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03




#27 AR-1254-2

R.T.: 6.021 min
Delta R.T.: 0.001 min
Response: 14065720
Conc: 25.04 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 7.332 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 6.433 min
Delta R.T.: 0.007 min
Response: 4098733
Conc: 4.20 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 7.616 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03




#29 AR-1254-4

R.T.: 6.598 min
Delta R.T.: -0.056 min
Response: 8696535
Conc: 12.17 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 7.494 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.598 min
Delta R.T.: 0.042 min
Response: 8696535
Conc: 13.28 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 7.749 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03




#32 AR-1260-2

R.T.: 6.741 min
Delta R.T.: -0.002 min
Response: 1225547
Conc: 1.34 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 9.357 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.175 min
Delta R.T.: 0.002 min
Response: 2710921
Conc: 1.64 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 10.257 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK03



#45 AR-1268-5

R.T.: 8.793 min
Delta R.T.: 0.002 min
Response: 3209700
Conc: 0.68 ng/ml