

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110222\
 Data File : PR057867.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2022 04:15
 Operator : AJ\MA
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 05:30:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 02 05:27:27 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.694	2.921	161.4E6	135.7E6	72.194	74.963
2) SA Decachlor...	9.645	8.168	160.3E6	223.6E6	71.044	73.402
Target Compounds						
16) L4 AR-1242-1	4.925	4.035	42120175	29254536	693.075	723.912
17) L4 AR-1242-2	4.946	4.053	57712441	46335753	703.065	720.044
18) L4 AR-1242-3	5.011	4.231	36538871	26178086	693.042	722.301
19) L4 AR-1242-4	5.115	4.323	30572593	20227899	706.931	701.828
20) L4 AR-1242-5	5.907	4.868	33951774	29243752	702.119	712.736



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110222\
Data File : PR057867.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Nov 2022 04:15
Operator : AJ\MA
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 05:30:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110222.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 02 05:27:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 161440259
Conc: 72.19 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 2.921 min
Delta R.T.: 0.000 min
Response: 135715046
Conc: 74.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.645 min
Delta R.T.: 0.002 min
Response: 160278538
Conc: 71.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.168 min
Delta R.T.: 0.000 min
Response: 223571009
Conc: 73.40 ng/ml



#16 AR-1242-1


R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 42120175
Conc: 693.07 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.035 min
Delta R.T.: 0.000 min
Response: 29254536
Conc: 723.91 ng/ml




#17 AR-1242-2

R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 57712441
Conc: 703.07 ng/ml



#17 AR-1242-2


R.T.: 4.053 min
Delta R.T.: 0.000 min
Response: 46335753
Conc: 720.04 ng/ml



#18 AR-1242-3


R.T.: 5.011 min
Delta R.T.: 0.000 min
Response: 36538871
Conc: 693.04 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 4.231 min
Delta R.T.: 0.000 min
Response: 26178086
Conc: 722.30 ng/ml




#19 AR-1242-4

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 30572593
Conc: 706.93 ng/ml



#19 AR-1242-4


R.T.: 4.323 min
Delta R.T.: 0.000 min
Response: 20227899
Conc: 701.83 ng/ml



#20 AR-1242-5

R.T.: 5.907 min
Delta R.T.: -0.001 min
Response: 33951774
Conc: 702.12 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 4.868 min
Delta R.T.: 0.000 min
Response: 29243752
Conc: 712.74 ng/ml