

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110222\
 Data File : PR057889.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2022 10:00
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR110222

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 10:34:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 02 09:30:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.694	2.920	115.1E6	93779975	53.556	55.583
2) SA Decachlor...	9.655	8.171	121.7E6	165.8E6	55.693	57.753
Target Compounds						
21) L5 AR-1248-1	4.925	4.035	24611662	16579004	551.232	562.298
22) L5 AR-1248-2	5.220	4.282	31167073	20022680	560.406	566.915
23) L5 AR-1248-3	5.434	4.323	37506136	21522891	567.431	562.050
24) L5 AR-1248-4	5.869	4.498	43055952	28392375	556.978	568.933
25) L5 AR-1248-5	5.909	4.910	42034343	31671974	560.953	563.243



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110222\
Data File : PR057889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Nov 2022 10:00
Operator : AJ\MA
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR110222

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 10:34:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110222.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 02 09:30:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 115057640
Conc: 53.56 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222




#1 Tetrachloro-m-xylene

R.T.: 2.920 min
Delta R.T.: 0.000 min
Response: 93779975
Conc: 55.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.655 min
Delta R.T.: 0.008 min
Response: 121675895
Conc: 55.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.171 min
Delta R.T.: 0.002 min
Response: 165813202
Conc: 57.75 ng/ml



#21 AR-1248-1


R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 24611662
Conc: 551.23 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222




#21 AR-1248-1

R.T.: 4.035 min
Delta R.T.: 0.000 min
Response: 16579004
Conc: 562.30 ng/ml




#22 AR-1248-2

R.T.: 5.220 min
Delta R.T.: 0.000 min
Response: 31167073
Conc: 560.41 ng/ml



#22 AR-1248-2


R.T.: 4.282 min
Delta R.T.: 0.000 min
Response: 20022680
Conc: 566.92 ng/ml



#23 AR-1248-3


R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 37506136
Conc: 567.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222




#23 AR-1248-3

R.T.: 4.323 min
Delta R.T.: 0.000 min
Response: 21522891
Conc: 562.05 ng/ml




#24 AR-1248-4

R.T.: 5.869 min
Delta R.T.: 0.000 min
Response: 43055952
Conc: 556.98 ng/ml



#24 AR-1248-4


R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 28392375
Conc: 568.93 ng/ml



#25 AR-1248-5

R.T.: 5.909 min
Delta R.T.: 0.000 min
Response: 42034343
Conc: 560.95 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222



#25 AR-1248-5

R.T.: 4.910 min
Delta R.T.: 0.000 min
Response: 31671974
Conc: 563.24 ng/ml