

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110222\
 Data File : PR057891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2022 10:31
 Operator : AJ\MA
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR110222

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 10:41:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 02 09:30:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.694	2.920	108.2E6	89490595	50.352	53.041
2)	SA Decachlor...	9.653	8.171	191.5E6	270.2E6	87.656	94.104
Target Compounds							
41)	L9 AR-1268-1	8.221	7.026	151.9E6	169.5E6	515.001	518.452
42)	L9 AR-1268-2	8.313	7.095	137.9E6	189.8E6	517.684	527.271
43)	L9 AR-1268-3	8.535	7.319	116.9E6	165.8E6	517.567	525.988
44)	L9 AR-1268-4	8.944	7.631	53782221	67415111	518.357	526.010
45)	L9 AR-1268-5	9.336	7.927	392.6E6	632.1E6	527.238	536.866



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110222\
Data File : PR057891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Nov 2022 10:31
Operator : AJ\MA
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR110222

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 10:41:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110222.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 02 09:30:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 108174496
Conc: 50.35 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222




#1 Tetrachloro-m-xylene

R.T.: 2.920 min
Delta R.T.: 0.000 min
Response: 89490595
Conc: 53.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.653 min
Delta R.T.: 0.005 min
Response: 191508341
Conc: 87.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.171 min
Delta R.T.: 0.002 min
Response: 270182897
Conc: 94.10 ng/ml



#41 AR-1268-1


R.T.: 8.221 min
Delta R.T.: 0.000 min
Response: 151860122
Conc: 515.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222




#41 AR-1268-1

R.T.: 7.026 min
Delta R.T.: 0.000 min
Response: 169478954
Conc: 518.45 ng/ml




#42 AR-1268-2

R.T.: 8.313 min
Delta R.T.: 0.000 min
Response: 137926419
Conc: 517.68 ng/ml



#42 AR-1268-2


R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 189811240
Conc: 527.27 ng/ml



#43 AR-1268-3


R.T.: 8.535 min
Delta R.T.: 0.000 min
Response: 116935858
Conc: 517.57 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222




#43 AR-1268-3

R.T.: 7.319 min
Delta R.T.: 0.000 min
Response: 165800166
Conc: 525.99 ng/ml




#44 AR-1268-4

R.T.: 8.944 min
Delta R.T.: 0.000 min
Response: 53782221
Conc: 518.36 ng/ml



#44 AR-1268-4


R.T.: 7.631 min
Delta R.T.: 0.000 min
Response: 67415111
Conc: 526.01 ng/ml



#45 AR-1268-5

R.T.: 9.336 min
Delta R.T.: 0.001 min
Response: 392644060
Conc: 527.24 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR110222



#45 AR-1268-5

R.T.: 7.927 min
Delta R.T.: 0.000 min
Response: 632137346
Conc: 536.87 ng/ml