

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR110320.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 04 04:25:33 2020
 Response Via : Initial Calibration

Calibration Files

750 =PR048434.D 1000 =PR048433.D 500 =PR048435.D
 250 =PR048436.D 50 =PR048437.D

	Compound	750	1000	500	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.540	2.479	2.543	2.511	2.167	2.448 E6	6.49
2) SA	Decachlorobiphenyl	2.484	2.362	2.603	2.839	2.887	2.635 E6	8.56
3) L1	AR-1016-1	0.908	0.855	0.936	1.011	0.962	0.934 E5	6.24
4) L1	AR-1016-2	1.276	1.227	1.326	1.370	1.327	1.305 E5	4.21
5) L1	AR-1016-3	7.762	7.334	8.099	8.583	8.009	7.957 E4	5.76
6) L1	AR-1016-4	6.448	6.144	6.757	6.941	6.276	6.513 E4	5.09
7) L1	AR-1016-5	6.293	5.992	6.594	6.924	6.112	6.383 E4	5.92
8) L2	AR-1221-1			2.684			2.684 E4	0.00
9) L2	AR-1221-2			1.985			1.985 E4	0.00
10) L2	AR-1221-3			6.110			6.110 E4	0.00
11) L3	AR-1232-1			5.350			5.350 E4	0.00
12) L3	AR-1232-2			2.775			2.775 E4	0.00
13) L3	AR-1232-3			5.471			5.471 E4	0.00
14) L3	AR-1232-4			2.663			2.663 E4	0.00
15) L3	AR-1232-5			2.064			2.064 E4	0.00
16) L4	AR-1242-1			6.399			6.399 E4	0.00
17) L4	AR-1242-2			9.104			9.104 E4	0.00
18) L4	AR-1242-3			5.566			5.566 E4	0.00
19) L4	AR-1242-4			4.563			4.563 E4	0.00
20) L4	AR-1242-5			4.997			4.997 E4	0.00
21) L5	AR-1248-1			4.814			4.814 E4	0.00
22) L5	AR-1248-2			6.746			6.746 E4	0.00
23) L5	AR-1248-3			7.498			7.498 E4	0.00
24) L5	AR-1248-4			8.706			8.706 E4	0.00
25) L5	AR-1248-5			8.387			8.387 E4	0.00
26) L6	AR-1254-1			8.777			8.777 E4	0.00
27) L6	AR-1254-2			1.323			1.323 E5	0.00
28) L6	AR-1254-3			1.399			1.399 E5	0.00
29) L6	AR-1254-4			1.083			1.083 E5	0.00
30) L6	AR-1254-5			1.152			1.152 E5	0.00
31) L7	AR-1260-1	1.082	1.027	1.134	1.194	1.245	1.136 E5	7.65
32) L7	AR-1260-2	1.327	1.265	1.392	1.479	1.556	1.404 E5	8.28
33) L7	AR-1260-3	1.030	0.980	1.079	1.131	1.112	1.066 E5	5.78
34) L7	AR-1260-4	1.199	1.153	1.258	1.307	1.295	1.242 E5	5.25
35) L7	AR-1260-5	2.463	2.383	2.502	2.588	2.576	2.502 E5	3.38
36) L8	AR-1262-1			1.352			1.352 E5	0.00
37) L8	AR-1262-2			2.476			2.476 E5	0.00
38) L8	AR-1262-3			1.744			1.744 E5	0.00
39) L8	AR-1262-4			1.346			1.346 E5	0.00
40) L8	AR-1262-5			9.979			9.979 E4	0.00
41) L9	AR-1268-1			3.233			3.233 E5	0.00
42) L9	AR-1268-2			3.050			3.050 E5	0.00
43) L9	AR-1268-3			2.556			2.556 E5	0.00
44) L9	AR-1268-4			1.188			1.188 E5	0.00
45) L9	AR-1268-5			8.528			8.528 E5	0.00

Signal #2 Calibration Files

750 =PR048434.D 1000 =PR048433.D 500 =PR048435.D
 250 =PR048436.D 50 =PR048437.D

	Compound	750	1000	500	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.838	2.687	2.911	3.080	3.002	2.904 E6	5.23
2) SA	Decachlorobiphenyl	1.472	1.428	1.509	1.500	1.394	1.461 E7	3.35
3) L1	AR-1016-1	7.838	7.536	8.609	8.514	8.196	8.139 E4	5.56
4) L1	AR-1016-2	1.667	1.568	1.700	1.819	1.589	1.669 E5	5.99
5) L1	AR-1016-3	1.005	0.958	1.047	1.060	0.902	0.995 E5	6.56
6) L1	AR-1016-4	4.537	4.310	4.821	4.968	3.693	4.466 E4	11.23

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR110320.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 04 04:25:33 2020
 Response Via : Initial Calibration

Calibration Files

750 =PR048434.D 1000 =PR048433.D 500 =PR048435.D
 250 =PR048436.D 50 =PR048437.D

Compound			750	1000	500	250	50	Avg	%RSD

7)	L1	AR-1016-5	7.808	7.396	8.328	8.435	9.154	8.224 E4	8.11
8)	L2	AR-1221-1			2.516			2.516 E4	0.00
9)	L2	AR-1221-2			2.192			2.192 E4	0.00
10)	L2	AR-1221-3			6.468			6.468 E4	0.00
11)	L3	AR-1232-1			5.522			5.522 E4	0.00
12)	L3	AR-1232-2			7.057			7.057 E4	0.00
13)	L3	AR-1232-3			4.343			4.343 E4	0.00
14)	L3	AR-1232-4			2.442			2.442 E4	0.00
15)	L3	AR-1232-5			2.845			2.845 E4	0.00
16)	L4	AR-1242-1			5.479			5.479 E4	0.00
17)	L4	AR-1242-2			1.164			1.164 E5	0.00
18)	L4	AR-1242-3			7.036			7.036 E4	0.00
19)	L4	AR-1242-4			4.600			4.600 E4	0.00
20)	L4	AR-1242-5			7.504			7.504 E4	0.00
21)	L5	AR-1248-1			4.389			4.389 E4	0.00
22)	L5	AR-1248-2			5.469			5.469 E4	0.00
23)	L5	AR-1248-3			6.686			6.686 E4	0.00
24)	L5	AR-1248-4			9.023			9.023 E4	0.00
25)	L5	AR-1248-5			1.380			1.380 E5	0.00
26)	L6	AR-1254-1			1.543			1.543 E5	0.00
27)	L6	AR-1254-2			1.690			1.690 E5	0.00
28)	L6	AR-1254-3			3.347			3.347 E5	0.00
29)	L6	AR-1254-4			5.507			5.507 E5	0.00
30)	L6	AR-1254-5			5.823			5.823 E5	0.00
31)	L7	AR-1260-1	1.790	1.696	1.882	2.102	2.033	1.900 E5	8.83
32)	L7	AR-1260-2	5.931	5.547	6.267	7.351	7.616	6.542 E5	13.77
33)	L7	AR-1260-3	3.738	3.543	3.952	4.559	5.568	4.272 E5	19.17
34)	L7	AR-1260-4	4.526	4.610	5.078	5.713	5.410	5.067 E5	10.04
35)	L7	AR-1260-5	2.119	2.064	2.136	2.119	1.739	2.035 E6	8.26
36)	L8	AR-1262-1			3.027			3.027 E5	0.00
37)	L8	AR-1262-2			2.051			2.051 E6	0.00
38)	L8	AR-1262-3			8.026			8.026 E5	0.00
39)	L8	AR-1262-4			1.367			1.367 E6	0.00
40)	L8	AR-1262-5			7.029			7.029 E5	0.00
41)	L9	AR-1268-1			2.614			2.614 E6	0.00
42)	L9	AR-1268-2			2.284			2.284 E6	0.00
43)	L9	AR-1268-3			1.961			1.961 E6	0.00
44)	L9	AR-1268-4			7.750			7.750 E5	0.00
45)	L9	AR-1268-5			5.643			5.643 E6	0.00

(#) = Out of Range