

## **CASE NARRATIVE**

**Garden State Laboratories, Inc.**

**Project Name: Waste Water 2020**

**Project # N/A**

**Chemtech Project # L4490**

**Test Name: VOC-TCLVOA-10**

### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 10/22/2020.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10 and VOCMS Group1. This data package contains results for VOC-TCLVOA-10.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOC-TCLVOA-10 was based on method 8260D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 15% in the Initial Calibration method (82X102120W.M) for Bromomethane, Methyl Iodide, Acrolein, are passing on Linear Regression, while Chloroethane is passing in Quadratic Regression.

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

### **E. Additional Comments:**

The temperature of the samples at the time of receipt was 13.2

Samples for MS/MSD for VOC analysis were not provided with this set of samples.

The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_