



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Tectonic Engineering**

**Project Name: PN20-1249 - 1111 Central Park Ave.**

**Project # N/A**

**Chemtech Project # L4510**

**Test Name: Hexavalent Chromium,Cyanide**

### **A. Number of Samples and Date of Receipt:**

27 Solid samples were received on 10/23/2020.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH\_NF, Herbicide Group1, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, PCB, PESTICIDE Group1, SVOCMS Group1, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS and VOCMS Group1. This data package contains results for Hexavalent Chromium,Cyanide.

### **C. Analytical Techniques:**

The analysis of Hexavalent Chromium was based on method 7196A and The analysis of Cyanide was based on method 9012B.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (CL-01-102320MS) analysis met criteria for all samples except for Hexavalent Chromium.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_